




**EDICT
OF
GOVERNMENT**

In order to promote public education and public safety, equal justice for all, a better informed citizenry, the rule of law, world trade and world peace, this legal document is hereby made available on a noncommercial basis, as it is the right of all humans to know and speak the laws that govern them.

EAS 80-7 (2006) (English): Butter
Methods of chemical analysis Part 7:
Determination of pH of the serum -
Potentiometric method

ISO INSIDE



BLANK PAGE





EAS 80-7:2006
ICS 67.100

EAST AFRICAN STANDARD

Butter — Methods of analysis — Part 7: Determination of pH of the serum — Potentiometric method

EAST AFRICAN COMMUNITY

Foreword

Development of the East African Standards has been necessitated by the need for harmonizing requirements governing quality of products and services in East Africa. It is envisaged that through harmonized standardization, trade barriers which are encountered when goods and services are exchanged within the Community will be removed.

In order to achieve this objective, the Partner States in the Community through their National Bureaux of Standards, have established an East African Standards Committee.

The Committee is composed of representatives of the National Standards Bodies in Partner States, together with the representatives from the private sectors and consumer organizations. Draft East African Standards are circulated to stakeholders through the National Standards Bodies in the Partner States. The comments received are discussed and incorporated before finalization of standards, in accordance with the procedures of the Community.

East African Standards are subject to review, to keep pace with technological advances. Users of the East African Standards are therefore expected to ensure that they always have the latest versions of the standards they are implementing.

© East African Community 2006 — All rights reserved*

East African Community

P O Box 1096

Arusha

Tanzania

Tel: 255 27 2504253/8

Fax: 255-27-2504481/2504255

E-Mail: eac@eachq.org

Web: www.each.org

* © 2006 EAC — All rights of exploitation in any form and by any means reserved worldwide for EAC Partner States' NSBs.

INTERNATIONAL
STANDARD

ISO
7238

IDF
104

Second edition
2004-12-01

**Butter — Determination of pH of the
serum — Potentiometric method**

*Beurre — Détermination du pH de la phase aqueuse — Méthode
potentiométrique*



Reference numbers
ISO 7238:2004(E)
IDF 104:2004(E)

© ISO and IDF 2004

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor the IDF accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IDF national committees. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat at the address given below.

© ISO and IDF 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO or IDF at the respective address below.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

International Dairy Federation
Diamant Building • Boulevard Auguste Reyers 80 • B-1030 Brussels
Tel. + 32 2 733 98 88
Fax + 32 2 733 04 13
E-mail info@fil-idf.org
Web www.fil-idf.org

Published in Switzerland

Contents

Foreword.....	iv
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	1
4 Principle	1
5 Reagents	1
6 Apparatus.....	2
7 Sampling	2
8 Procedure.....	3
8.1 Test portion	3
8.2 Separation of the serum	3
8.3 Calibration of the pH-meter.....	3
8.4 Determination	3
8.5 Cleaning the electrodes.....	3
9 Expression of results.....	3
10 Precision	4
10.1 Repeatability	4
10.2 Reproducibility	4
11 Test report.....	4
Bibliography	5

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7238|IDF 104 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

This edition of ISO 7238|IDF 104 cancels and replaces ISO 7238:1983, of which it constitutes a minor revision. Only editorial changes have been made.

Foreword

IDF (the International Dairy Federation) is a worldwide federation of the dairy sector with a National Committee in every member country. Every National Committee has the right to be represented on the IDF Standing Committees carrying out the technical work. IDF collaborates with ISO and AOAC International in the development of standard methods of analysis and sampling for milk and milk products.

Draft International Standards adopted by the Action Teams and Standing Committees are circulated to the National Committees for voting. Publication as an International Standard requires approval by at least 50 % of the National Committees casting a vote.

ISO 7238|IDF 104 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

All work was carried out by the Joint ISO/IDF/AOAC Group of Experts, *pH of butter* (E36), under the aegis of its project leader, Mr L.J. Pootvliet (NL).

This edition of ISO 7238|IDF 104 cancels and replaces IDF 104A:1984. Only editorial changes have been made.

Butter — Determination of pH of the serum — Potentiometric method

1 Scope

This International Standard specifies a potentiometric method for the determination of the pH of the serum from all types of butter.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

pH of butter serum

potential difference at the measuring temperature between two electrodes immersed in butter serum, determined by the procedure specified in this International Standard

NOTE It is expressed in pH units.

4 Principle

The potential difference is measured between a glass electrode and a reference electrode in the serum separated from melted butter.

5 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified, and recently distilled water that has been protected from absorption of carbon dioxide and that complies with the requirements for grade 1 water specified in ISO 3696:1987.

5.1 Buffer solutions, for calibration of the pH-meter.

Two standard buffer solutions, having pH values known to the second decimal place at the measuring temperature, and which will bracket the pH value of the serum obtained from the test portion, shall be used, for example a buffer solution of pH approximately 4 and another of pH approximately 7.

EXAMPLES The following buffer solutions may be used:

a) Buffer solution of pH 4,00 at 20 °C and 4,01 at 25 °C

Dissolve, in water, 10,12 g of potassium hydrogen phthalate ($\text{KHC}_8\text{H}_4\text{O}_4$), which has been previously dried to constant mass at 120 °C. Make up to 1 000 ml with water at the measuring temperature and mix well.

Preserve the solution by adding approximately 2 ml of chloroform or carbon tetrachloride.

b) Buffer solution of pH 6,88 at 20 °C and 6,86 at 25 °C

Dissolve, in water, 3,388 g of potassium dihydrogen orthophosphate (KH_2PO_4) and 3,533 g of disodium hydrogen orthophosphate (Na_2HPO_4), both compounds having been previously dried to constant mass at 120 °C. Make up to 1 000 ml with water at the measuring temperature and mix well.

Preserve the solution by adding approximately 2 ml of chloroform or carbon tetrachloride.

6 Apparatus

Usual laboratory apparatus and, in particular, the following.

6.1 pH-meter, of minimum sensitivity 0,01 pH unit, with a glass electrode and a suitable reference electrode, and with temperature compensation.

The glass and reference electrodes may be assembled into a system of combined electrodes.

6.2 Centrifuge (if required), of the vertical-loading type¹⁾, capable of attaining a relative radial acceleration of approximately 375 g.

6.3 Centrifuge tubes (if required), of capacity approximately 50 ml, with suitable stoppers.

6.4 Test tubes, of capacity approximately 12 ml, internal diameter 16 mm to 20 mm.

6.5 Water bath (if required), capable of being maintained at 65 °C.

6.6 Ice-water bath (if required).

6.7 Calibrated thermometer, accurate to 1 °C.

7 Sampling

A representative sample should have been sent to the laboratory. It should not have been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 707 [1].

1) For details of a suitable centrifuge, reference should be made to ISO 2446:1976, 6.6.

8 Procedure

8.1 Test portion

Take approximately 50 g of the sample.

8.2 Separation of the serum

Separate the serum from the butter by an appropriate method (see the Note). Transfer the serum (including the protein) to a test tube (6.4), and bring it to the measuring temperature.

NOTE There are many methods for the separation of the serum from the butter, one of which is as follows.

Transfer the test portion to a centrifuge tube (6.3), and place the centrifuge tube in the water bath (6.5). Two layers will be formed by the melting butter.

As soon as the butter has melted (after 3 min to 5 min), close the centrifuge tube with the stopper, place it, with the stoppered end downwards, in the tube holder and centrifuge for 5 min at a relative radial acceleration of approximately 375 *g*.

Immediately immerse the centrifuge tube, with the stoppered end downwards, in the ice-water bath (6.6) and leave until the fat has completely congealed.

8.3 Calibration of the pH-meter

Adjust the temperature of the buffer solutions (5.1) to the measuring temperature and calibrate the pH-meter in accordance with the manufacturer's instructions.

If a series of samples is being tested, check the calibration of the pH-meter with one or both buffer solutions at least every 30 min.

8.4 Determination

Introduce the electrodes into the serum at the measuring temperature.

Carry out the determination using the procedure appropriate to the pH-meter used. When the reading becomes constant, read the pH directly from the scale of the instrument, to the nearest 0,01 pH unit.

Introduce the thermometer (6.7) into the serum and read the measuring temperature.

8.5 Cleaning the electrodes

Clean the electrodes by rinsing consecutively with acetone at room temperature and water at 30 °C to 35 °C.

Dab them dry with a clean paper tissue.

9 Expression of results

Record the measured pH to the nearest 0,01 pH unit, together with the measuring temperature.

10 Precision

10.1 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than 0,03 pH unit.

10.2 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than 0,10 pH unit.

11 Test report

The test report shall specify:

- a) all information necessary for the complete identification of the sample;
- b) the sampling method used, if known;
- c) the test method used, with reference to this International Standard;
- d) all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test results;
- e) the test results obtained, or, if the repeatability has been checked, the final quoted result obtained.

Bibliography

- [1] ISO 707, *Milk and milk products — Guidance on sampling* ²⁾
- [2] ISO 2446:1976, *Milk — Determination of fat content (Routine method)*

2) Equivalent to IDF 50.

ISO 7238:2004(E)
IDF 104:2004(E)

ICS 67.100.20

Price based on 5 pages

