

इंटरनेट

मानक

### Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 9377 (1979): Specification for apparatus for aggregate impact value [CED 2: Cement and Concrete]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”



BLANK PAGE



*Indian Standard*

**SPECIFICATION FOR  
APPARATUS FOR AGGREGATE  
IMPACT VALUE**

( Second Reprint MARCH 1996 )

UDC 620.178.153.2.05:666.972.12

© *Copyright* 1980

**BUREAU OF INDIAN STANDARDS**  
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG  
NEW DELHI 110002

# *Indian Standard*

## SPECIFICATION FOR APPARATUS FOR AGGREGATE IMPACT VALUE

Cement and Concrete Sectional Committee, BDC 2

### *Chairman*

DR H. C. VISVESVARAYA

### *Representing*

Cement Research Institute of India, New Delhi

### *Members*

ADDITIONAL DIRECTOR, STANDARDS ( B & S )	Research, Designs & Standards Organization ( Ministry of Railways ), Lucknow
DEPUTY DIRECTOR, STANDARDS ( B & S ) ( <i>Alternate</i> )	
SHRI K. C. AGGARWAL	Hindustan Prefab Ltd, New Delhi
SHRI C. L. KASLIWAL ( <i>Alternate</i> )	
SHRI K. P. BANERJEE	Larsen & Toubro Ltd, Bombay
SHRI HARISH N. MALANI ( <i>Alternate</i> )	
SHRI S. K. BANERJEE	National Test House, Calcutta
SHRI R. N. BANSAL	Beas Designs Organization, Nangal Township
SHRI T. C. GARG ( <i>Alternate</i> )	
CHIEF ENGINEER ( DESIGNS )	Central Public Works Department, New Delhi
EXECUTIVE ENGINEER ( DESIGNS ) III ( <i>Alternate</i> )	
CHIEF ENGINEER ( PROJECTS )	Irrigation Department, Government of Punjab
DIRECTOR, IPRI ( <i>Alternate</i> )	
DIRECTOR ( CSMRS )	Central Water Commission, New Delhi
DEPUTY DIRECTOR ( CSMRS ) ( <i>Alternate</i> )	
DR R. K. GHOSH	Central Road Research Institute ( CSIR ), New Delhi
SHRI Y. R. PHULL ( <i>Alternate I</i> )	
SHRI M. DINAKARAN ( <i>Alternate II</i> )	
DR R. K. GHOSH	Indian Roads Congress, New Delhi

( Continued on page 2 )

© Copyright 1980

BUREAU OF INDIAN STANDARDS

This publication is protected under the *Indian Copyright Act* ( XIV of 1957 ) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

( Continued from page 1 )

<i>Members</i>	<i>Representing</i>
SHRI B. R. GOVIND	Engineer-in-Chief's Branch, Army Headquarters
SHRI P. C. JAIN ( <i>Alternate</i> )	
SHRI A. K. GUPTA	Hyderabad Asbestos Cement Products Ltd, Hyderabad
DR R. R. HATTIANGADI	The Associated Cement Companies Ltd, Bombay
SHRI P. J. JAGUS ( <i>Alternate</i> )	
DR IQBAL ALI	Engineering Research Laboratories, Hyderabad
SHRI S. R. KULKARNI	M. N. Dastur & Co ( Pvt ) Ltd, Calcutta
SHRI S. K. LAHA	The Institution of Engineers ( India ), Calcutta
SHRI B. T. UNWALLA ( <i>Alternate</i> )	
DR MOHAN RAI	Central Building Research Institute ( CSIR ), Roorkee
DR S. S. REHSI ( <i>Alternate</i> )	
SHRI K. K. NAMBIAR	In personal capacity ( 'Ramanalaya' 11 First Crescent Park Road, Gandhinagar, Adyar, Madras )
DR M. RAMAIAH	Structural Engineering Research Centre ( CSIR ), Roorkee
DR N. S. BHAL ( <i>Alternate</i> )	
SHRI G. RAMDAS	Directorate General of Supplies & Disposals, New Delhi
DR A. V. R. RAO	National Buildings Organization, New Delhi
SHRI J. SEN GUPTA ( <i>Alternate</i> )	
SHRI R. V. CHALAPATHI RAO	Geological Survey of India, Calcutta
SHRI S. ROY ( <i>Alternate</i> )	
SHRI T. N. S. RAO	Gammon India Ltd, Bombay
SHRI S. R. PINHEIRO ( <i>Alternate</i> )	
SHRI ARJUN RIJHSINGHANI	Cement Corporation of India Ltd, New Delhi
SHRI K. VITHAL RAO ( <i>Alternate</i> )	
SECRETARY	Central Board of Irrigation and Power, New Delhi
DEPUTY SECRETARY ( I ) ( <i>Alternate</i> )	
SHRI N. SIVAGURU	Roads Wing, Ministry of Shipping and Transport
SHRI R. L. KAPOOR ( <i>Alternate</i> )	
SHRI K. A. SUBRAMANIAM	The India Cements Ltd, Madras
SHRI P. S. RAMACHANDRAN ( <i>Alternate</i> )	
SUPERINTENDING ENGINEER	Public Works Department, Government of Tamil ( DESIGNS ) Nadu
EXECUTIVE ENGINEER	
( SM & R DIVISION ) ( <i>Alternate</i> )	
SHRI L. SWAROOP	Dalmia Cement ( Bharat ) Ltd, New Delhi
SHRI A. V. RAMANA ( <i>Alternate</i> )	
SHRI B. T. UNWALLA	The Concrete Association of India, Bombay
SHRI Y. K. MEHTA ( <i>Alternate</i> )	
SHRI D. AJITHA SIMHA,	Director General, ISI ( <i>Ex-officio Member</i> ) Director ( Civ Engg )

*Secretary*

SHRI M. N. NEELAKANDHAN  
Assistant Director ( Civ Engg ), ISI

( Continued on page 7 )

# *Indian Standard*

## SPECIFICATION FOR APPARATUS FOR AGGREGATE IMPACT VALUE

### 0. FOREWORD

**0.1** This Indian Standard was adopted by the Indian Standards Institution on 20 December 1979, after the draft finalized by the Cement and Concrete Sectional Committee had been approved by the Civil Engineering Division Council.

**0.2** The Indian Standards Institution has already published a series of standards on methods of testing cement and concrete. It has been recognized that reliable and intercomparable test results could be obtained only with standard testing equipment capable of giving the desired level of accuracy. The Sectional Committee has therefore proposed to bring out a series of specifications covering the requirements of equipment used for testing cement and concrete, to encourage their development and manufacture in the country.

**0.3** This standard specifies the requirements for the apparatus used for determining the aggregate impact value of coarse aggregate. The aggregate impact value gives a relative measure of the resistance of an aggregate to sudden shock or impact, which in some aggregates differs from its resistance to a slow compressive load. The method of determining aggregate impact value has been covered in IS : 2386 ( Part IV )-1963\*.

**0.4** In the formulation of this standard, due weightage has been given to international co-ordination among the standards and practices prevailing in different countries in addition to relating it to the practices in the field in this country.

**0.5** For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS : 2-1960†. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

---

\*Methods of test for aggregates for concrete: Part IV Mechanical properties.

†Rules for rounding off numerical values ( *revised* ).

## 1. SCOPE

**1.1** This standard covers the requirements of apparatus for determining the impact value of coarse aggregates.

## 2. APPARATUS

**2.1** The apparatus shall consist of an impact testing machine and other accessories ( see 3 and 4 ).

## 3. IMPACT TESTING MACHINE

**3.1** The general form and salient dimensions of machine shall be as described in Fig. 1 and 3.2 to 3.2.2.

**Note** — A stroke counter of suitable design may be provided to automatically record the number of blows.

**3.2** The total mass of the machine shall be between 45 to 60 kg. The machine shall have a cast iron base weighing between 20 to 30 kg with a plane lower surface of not less than 300 mm diameter, and supported on a plane concrete block of 600 × 600 × 450 mm size. The machine shall be prevented from rocking by securing it firmly and evenly to the foundation. Three 10 mm diameter holes shall be provided in the base plate for this purpose. Alternatively the machine may be supported on a plane metal plate cast into the foundation.

**3.2.1 Steel Cup** — A cylindrical mild steel cup of the following internal dimensions shall be provided at the centre of the base of the machine in a manner such that it can be easily and rigidly fastened to and removed from the base:

Diameter	$100 \begin{matrix} + 0.25 \\ - 0 \end{matrix}$ mm
Depth	$50 \pm 0.25$ mm
Thickness	$7 \pm 1$ mm

A chamfer of 2 to 3 mm shall be provided at the inner lip of the steel cup. The inner surface of the cup shall be case hardened and shall have a hardness of not less than 650 VH or equivalent.

**3.2.2 Steel Hammer** — A steel tup or hammer, the lower end of which shall be cylindrical in shape with 2 to 3 mm chamfer at the lower edge. The hammer shall slide freely between the vertical steel guides so arranged that cylindrical lower end of the hammer is above and concentric with steel cup. Arrangement shall be provided for lifting the hammer and allowing it to fall freely between the vertical guides for a height of  $380 \pm 5$  mm on the test sample in the cup. It shall be possible to adjust height of fall within 5 mm. Arrangement shall also be provided for locking the hammer while fastening or removing the cup.



The release mechanism for hammer shall be quick release type giving reproducible height of drop between successive drops within  $\pm 0.5$  mm. The mass of the hammer and other dimensions shall be as below:

Mass of hammer	$13.75 \pm 0.25$ kg
Diameter of cylindrical lower end of hammer	$98 \begin{smallmatrix} + 0 \\ - 1 \end{smallmatrix}$ mm
Length of the cylindrical lower end	$50 \pm 1$ mm
Diameter of shank of hammer	$75 \pm 1$ mm

The lower surface of the hammer shall be case hardened and shall have a hardness not less than 650 VH or equivalent.

#### 4. ACCESSORIES

**4.1 Cylindrical Measures** — A cylindrical steel measure tared to the nearest gram, of sufficient rigidity to retain its form under rough usage, and of the following dimensions:

Diameter	$75 \pm 1$ mm
Depth	$50 \pm 1$ mm
Shell thickness, <i>Min</i>	5 mm

**4.2 Tamping Rod** — A straight steel tamping rod of 10 mm diameter, 230 mm long and rounded at one end.


#### 5. MARKING

**5.1** The following information shall be clearly and indelibly marked on each component of the apparatus in a way that it does not interfere with the performance of the apparatus:

- a) Name of the manufacturer or his registered trade-mark or both, and
- b) Date of manufacture.

**5.1.1** The apparatus may also be marked with the ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution ( Certification Marks ) Act and the Rules and Regulations made thereunder. The ISI Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products are also continuously checked by ISI for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.



All dimensions in millimetres.

FIG. 1 AGGREGATE IMPACT TEST MACHINE

( Continued from page 2 )

# Instruments for Cement and Concrete Testing Subcommittee, BDC 2 : 10

*Convener*

DR IQBAL ALI

*Representing*

Engineering Research Laboratories, Hyderabad

*Members*

PROF B. M. AHUJA

SHRI T. P. EKAMBARAM

DR R. K. GHOSH

Indian Institute of Technology, New Delhi

Highways Research Station, Madras

Central Road Research Institute ( CSIR ),  
New Delhi

SHRI K. L. SETHI ( Alternate )

SHRI H. K. GUHA

All India Instruments Manufacturers and Dealers  
Association, Bombay

SHRI V. K. VASUDEVAN ( Alternate )

SHRI P. J. JAGUS

The Associated Cement Companies Ltd, Bombay

SHRI D. A. WADIA ( Alternate )

SHRI M. R. JOSHI

Research & Development Organization ( Ministry  
of Defence ), Pune

SHRI Y. P. PATHAK ( Alternate )

SHRI E. K. RAMACHANDRAN

National Test House, Calcutta

PROF C. K. RAMESH

Indian Institute of Technology, Bombay

DR R. S. AYYAR ( Alternate )

SHRI M. V. RANGA RAO

Cement Research Institute of India, New Delhi

DR K. C. NARANG ( Alternate )

DR S. S. REHSI

Central Building Research Institute ( CSIR ),  
Roorkee

SHRI J. P. KAUSHISH ( Alternate )

SHRI M. M. D. SETH

Public Works Department, Government of Uttar  
Pradesh

SHRI J. P. BHATNAGAR ( Alternate )

SHRI H. C. VERMA

Associated Instrument Manufacturers ( India )  
Private Ltd, New Delhi

SHRI A. V. SHASTRI ( Alternate )

# BUREAU OF INDIAN STANDARDS

## Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002

Telephones: 331 01 31, 331 13 75

Telegrams: Manaksanstha  
( Common to all Offices )

## Regional Offices:

## Telephone

Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg,  
NEW DELHI 110002

{ 331 01 31  
{ 331 13 75

\*Eastern : 1/14 C. I. T. Scheme VII M, V. I. P. Road,  
Manikotla, CALCUTTA 700054

36 24 99

Northern : SCO 445-446, Sector 35-C,  
CHANDIGARH 160036

{ 2 18 43  
{ 3 16 41

Southern : C. I. T. Campus, MADRAS 600113

{ 41 24 42  
{ 41 25 19

†Western : Manakalaya, E9 MIDC, Marol, Andheri ( East ),  
BOMBAY 400093

41 29 16

6 32 92 95

## Branch Offices:

'Pushpak', Nurmohamed Shaikh Marg, Khanpur,  
AHMADABAD 380001

{ 2 63 48  
{ 2 63 49

‡Peenya Industrial Area 1st Stage, Bangalore Tumkur Road  
BANGALORE 560058

{ 38 49 55  
{ 38 49 56

Gangotri Complex, 5th Floor, Bhadbhada Road, T. T. Nagar,  
BHOPAL 462003

6 67 16

Plot No. 82/83, Lewis Road, BHUBANESHWAR 751002

5 36 27

53/5, Ward No. 29, R.G. Barua Road, 5th Byelane,  
GUWAHATI 781003

3 31 77

5-8-56C L. N. Gupta Marg ( Nampally Station Road ),  
HYDERABAD 500001

23 10 83

R14 Yudhister Marg, C Scheme, JAIPUR 302005

{ 5 34 71  
{ 6 98 32

117/418 B Sarvodaya Nagar, KANPUR 208005

{ 21 68 76  
{ 21 82 92

Patliputra Industrial Estate, PATNA 800013

6 23 05

T.C. No. 14/1421, University P.O., Palayam

{ 6 21 04

TRIVANDRUM 695035

{ 6 21 17

## Inspection Offices ( With Sale Point ):

Pushpanjali, First Floor, 205-A West High Court Road,  
Shankar Nagar Square, NAGPUR 440010

2 51 71

Institution of Engineers ( India ) Building, 1332 Shivaji Nagar,  
PUNE 411005

5 24 35

\*Sales Office in Calcutta is at 5 Chowringhee Approach, P. O. Princep 27 68 00  
Street, Calcutta 700072

†Sales Office in Bombay is at Novelty Chambers, Grant Road, 89 65 28  
Bombay 400007

‡Sales Office in Bangalore is at Unity Building, Narasimharaja Square, 22 36 71  
Bangalore 560002

Reprography Unit, BIS, New Delhi, India