

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 12504-1 (1988): Tracheal Tubes, Part 1: General Requirements [MHD 13: Veterinary Hospital Planning and Surgical Instruments]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



Indian Standard

SPECIFICATION FOR TRACHEAL TUBES

PART 1 GENERAL REQUIREMENTS

भारतीय मानक

श्वंस नलिकाओं की विशिष्टि

भाग 1 सामान्य अपेक्षाएं

UDC 616·231-089·5 : 621·643·2

**DESCRIPTORS : MEDICAL EQUIPMENT. ANAESTHESIA. ENDOTRACHEAL TUBES.
SPECIFICATIONS. DESIGNATION. MARKING**

© BIS 1989

**BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002**

Indian Standard

SPECIFICATION FOR TRACHEAL TUBES

PART 1 GENERAL REQUIREMENTS

NATIONAL FOREWORD

This Indian Standard (Part 1), which is identical with **ISO 5361-1:1988** Tracheal tubes — Part 1: General requirements', issued by the International Organization for Standardization (**ISO**), was adopted by the Bureau of Indian Standards on 25 April 1988 on the recommendation of the **Anaesthetic**, Resuscitation and Allied Equipment Sectional Committee (CPDC 13) and approval of the Consumer Products and Medical instruments Division Council.

The text of **ISO** standard has been approved as suitable for publication as Indian Standard without deviations. Certain conventions are, however, not identical to those used in Indian Standards. Attention is particularly drawn to the following:

- a) Wherever the words 'international Standard' appear referring to this standard, they should be read as 'Indian Standard'.
- b) Comma (,) has been used as a decimal marker while in Indian Standards, the current practice is to use a point (.) as the decimal marker.

This Indian Standard is in five parts. The other four parts are:

- Part 2 Oro-tracheal and naso-tracheal tubes of the Magiii type (plain and cuffed),
- Part 3 Murphy type,
- Part 4 Cole type, and
- Part 5 Requirements and methods of test for cuffs and tubes.

CROSS REFERENCE

in this Indian Standard, the following international Standard is referred to, Read in its **place the** following:

<i>International Standard</i>	<i>Indian Standard</i>	<i>Degree of Correspondence</i>
ISO 5361/4 : 1984 Tracheal tubes — Part 4 : Cole type	IS 12504 (Part 4): 1988 Specification for tracheal tubes: Part 4 Cole type	Identical

The technical committee responsible for the Preparation Of this Indian Standard has reviewed the provisions of the following **ISO** standard and has decided that it is acceptable for use in **conjunc-**tion with this standard:

ISO 4135:1979 Anaesthesiology — Vocabulary

As in the Original Standard, this Page is Intentionally Left Blank

0 Introduction

This International Standard is one of a series dealing with anaesthetic equipment and medical breathing machines, and specifies the basic requirements for the more commonly used types of tracheal tubes, whether made of rubber or of other elastomeric material. Tubes with walls reinforced with metal or nylon, tubes with shoulders (except for those in ISO 5361-4), tapering tubes, or the many types of special tubes devised for use in thoracic surgery, are not specifically covered although most may have the size designated by the inside diameter as required by this specification.

The inside diameter has been specified for the purpose of size designation.

1 Scope and field of application

This part of ISO 5361 specifies general requirements for tracheal (endotracheal) tubes.

2 References

ISO 4135, *Anaesthesiology — Vocabulary.*

ISO 5361-4, *Tracheal tubes — Part 4: Cole type.*

3 Definitions

Definitions are taken from ISO 4135, except where marked with an asterisk.

3.1 tracheal tube (endotracheal tube): Tube designed for insertion through the larynx into the trachea to convey gases and vapours to and from the trachea.

3.2 oro-tracheal tube: Tracheal tube for insertion through the mouth into the trachea.

3.3 naso-tracheal tube: Tracheal tube for insertion through the nose into the trachea.

3.4 patient end: That end of the tracheal tube which is intended to be inserted into the trachea.

3.5 machine end: That end of the tracheal tube which is intended to project from the patient.

3.6 bevel: Slanted portion at the patient end of the tracheal tube.

3.7 angle of bevel*: Acute angle between the plane of the bevel and the longitudinal axis of the tracheal tube at the patient end.

3.6 cuff*: Inflatable sleeve fitted near the patient end of the tracheal tube to provide an effective fit between the tube and the trachea.

3.9 bonded cuff: Cuff permanently attached to the tracheal tube.

3.10 inflating tube: Tube through which the cuff is inflated.

3.11 pilot balloon*: Balloon fitted to the inflating tube to indicate inflation of the cuff.

4 Size designation

The size of tracheal tubes shall be designated by a number giving the nominal inside diameter expressed in millimetres.

5 Size range

The size range 2.5 to 11 (except for Cole tubes specified in ISO 5361-4) of tracheal tubes shall be based on sizes having the nominal inside diameter of a whole number of millimetres or of the intermediate 0.5 mm steps.

6 Materials

6.1 Tracheal tubes, including cuffs, in their ready-for-use state shall be compatible with the human tissues with which they are intended to be used as determined by the implantation test given in the annex.

6.2 Unless designated for single use, tracheal tubes should be reasonably resistant to deterioration by methods of cleaning disinfection and sterilization as recommended by the manufacturer or the supplier. Such tubes should withstand accepted methods of steam sterilization.

The recommended method or methods of sterilization should not produce changes in the tube material which will render the tracheal tube incompatible with human tissues with which it is **intended** to be used (see also 6.1).

6.3 Tracheal tubes under normal conditions of use should be reasonably resistant to deterioration by anaesthetic vapours and gases.

6.4 The material used for the manufacture of the tubes shall have sufficient rigidity to allow the construction of a tube with the thinnest possible wall which, at the same time, maintains resistance to kinking. When in place it shall be flexible and soft enough to conform to the patient's anatomy without exerting undue pressure on the body tissues.

6.5 Tracheal tubes should be readily detectable by X-ray either by the nature of the material of which they are made or by the provision of a marker at the patient end.

7 Lumen

The lumen of the tracheal tube shall be essentially circular in a plane at right angles to the long axis.

8 Finish

Tracheal tubes shall be smooth on the outside and inside. ~~The~~ cuffs shall have a smooth surface.

9 Marking

9.1 Marking materials used on tracheal tubes shall be of a contrasting **colour**, shall be compatible with the human tissues with which the tubes are intended to be used, shall resist anaesthetic agents and, unless designated for single use, shall resist agents and methods commonly used for cleaning and sterilizing (see also clause 6).

9.2 The marking of tracheal tubes shall be durable and legible and shall include the following:

- a) the name or trade-mark of the manufacturer or supplier;
- b) the size designation in accordance with clause 4;
- c) name of country of origin, if required;
- d) tracheal tubes intended for single use shall be clearly marked to indicate single use or marked in accordance with the requirements of the relevant national regulatory authorities;
- e) depth (length) mark(s) in centimetres measured from the patient end. This requirement is optional for Cole tubes.

10 Information to be supplied by manufacturer

The manufacturer shall provide a warning with the tracheal tube that cuff volume or pressure can either increase or decrease owing to diffusion of nitrous oxide/oxygen mixtures or air, and shall include recommendations for minimizing this effect.

A N N E X

Implantation test¹⁾

(This annex forms an integral part of the Standard)

A.1 General

The implantation test is designed for the evaluation of a plastics material in direct contact with living tissue. Care shall be taken in the preparation of the implant strips and their proper implantation under aseptic conditions.

A.2 Preparation of test samples

Prepare for implantation eight strips of the sample and four strips of **USP Negative Control Plastic RS²⁾**. Each strip shall measure not less than 10 mm x 1 mm. The edges of the strips should be as smooth as possible to avoid additional mechanical trauma upon implantation. Strips of the specified minimum size shall be implanted by means of a hypodermic needle such as a 15 gauge needle with intravenous point and of 19 mm (0.75 inch) **can-**nula length, and a sterile trocar. Use either presterilized needles into which the sterile plastics strips are aseptically inserted, or insert each clean strip into a needle, the **can-**nula and hub of which are protected with an appropriate cover, and then subjected to the appropriate sterilization procedure.

NOTE -Allowance should be made for proper **de-**gassing if agents such as ethylene oxide are used.

A.3 Test animal

Select healthy, adult rabbits weighing not less than 2.5 kg, and whose paravertebral muscles are sufficiently large in size to allow for implantation of the test strips. Do not use any muscular tissue other than the paravertebral site. The animals may be anaesthetized with

a commonly used anaesthetic agent to a degree deep enough to prevent muscular movements, such as twitching.

A.4 Procedure

Perform the test in a clean area. On the day of the test or up to 20 h before testing, clip the fur of the animals on both sides of the spinal column. Remove loose hair by means of vacuum.

Implant four strips of the sample into the paravertebral muscle on one side of the spine of each of two rabbits, 2.5 to 5 cm from the mid-line and parallel to the spinal column, and about 2.5 cm apart from each other. In a similar fashion implant two strips of USP Negative Control Plastic RS in the opposite muscle of each animal. Insert a **stylet** into the needle to hold the plastics strip in the tissue while withdrawing the needle. If excessive bleeding is observed after implantation of a strip, place a duplicate strip at another site. Close the incision after implantation is complete.

Keep the animals for a period of not less than 72 h and sacrifice them at the end of the observation period by administering an overdose of an anaesthetic agent. Allow sufficient time to elapse for the tissue to be cut without bleeding. Examine macroscopically the area of the tissue surrounding the centre portion of each implant strip. Use a magnifying lens if necessary. The tissue immediately surrounding the USP Negative Control Plastic RS strips appears normal and entirely free from hemorrhage, film, or encapsulation. The requirements of the test are met if, in each rabbit, the reaction to not more than one of the four sample strips is significantly greater than that to the strips of USP Negative Control Plastic RS.

1) Taken from US Pharmacopeia.

2) USP Negative Control Plastic RS is a trade name for a commercial product available from the US Pharmacopeial Convention *Inc.*, 12601 Twinbrook Parkway, Rockville, Maryland 20852, USA. At present, no other products intended for this purpose are known to be available commercially. This information is given for the convenience of the users of this International Standard and does not constitute an endorsement of this product by **ISO**.

Standard Mark

The use of the Standard Mark is governed by the provisions of the *Bureau of Indian Standards Act, 1986* and the Rules and Regulations made thereunder. The Standard Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well defined system of inspection, testing and quality control which is devised and supervised by BIS and operated by the producer. Standard marked products are also continuously checked by BIS for conformity to that standard as a further safeguard. Details of conditions under which a **licence** for the use of the Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 1986* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of **BIS**. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Revision of Indian Standards

Indian Standards are reviewed periodically and revised, when necessary and amendments, if any, are issued from time to time. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition. Comments on this Indian Standard may be sent to BIS giving the following reference:

Doc : No. CPDC 13 (1987)

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002
Telephones : 331 01 31, 331 13 75

Telegrams : Manaksanstha
(Common to all Offices)

Regional Offices:

Telephone

Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg,
NEW DELHI 110002

{ 331 331 01 13 3175

Eastern : **1/14 C. I. T.** Scheme VII M, V. I. P. Road, Maniktola,
CALCUTTA 700054

36 24 99

Northern : **SCO** 445-446, Sector 35-C, **CHANDIGARH** 160036

{ 32 18 16 41 43

Southern : **C. I. T.** Campus, IV Cross Road, Madras 600113

{ 41 24 42
41 25 19
41 29 16

Western : **Manakalaya, E9** MIDC, Marol, Andheri (East)
BOMBAY

6 32 92 95

Branches : AHMADABAD. BANGALORE. BHOPAL. BHUBANESHWAR.
GUWAHATI. HYDERABAD. JAIPUR. KANPUR. PATNA.
TRIVANDRUM.