

इंटरनेट

मानक

### Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 1448-32 (1992): Methods of test for petroleum and its products, Part 32: Density and relative density [PCD 1: Methods of Measurement and Test for Petroleum, Petroleum Products and Lubricants]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”



BLANK PAGE



भारतीय मानक

पेट्रोलियम और इसके उत्पाद — परीक्षण पद्धतियाँ

[ पी : 32 ]

घनत्व और आपेक्षिक घनत्व

( दूसरा पुनरीक्षण )

*Indian Standard*

PETROLEUM AND ITS PRODUCTS —  
METHODS OF TEST

[ P : 32 ]

DENSITY AND RELATIVE DENSITY

*( Second Revision )*

UDC 665-61-7 : 531-756

© BIS 1992

BUREAU OF INDIAN STANDARDS  
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG  
NEW DELHI 110002

## FOREWORD

This Indian Standard [ P : 32 ] ( Second Revision ) was adopted by the Bureau of Indian Standards, after the draft finalized by the Methods of Test for Petroleum, Petroleum Products and Lubricants Sectional Committee had been approved by the Petroleum, Coal and Related Products Division Council.

While preparing the standard considerable assistance has been derived from the standards IP 59/82, IP 189/86 and IP 190/86, issued by the Institute of Petroleum, London ( U. K. ).

Provisions have been made in this standard for the conversion of density and relative density from the test temperature to 15°C or 20°C by using the ASTM/IP Petroleum Measurement Tables ( Clause 3.4.1 ). Accordingly, use of various tables viz. Tables 23A, 23B, 23D, 24A, 24B, 24D, 53A, 53B, 53D, 54A, 54B, and 54D of IP 200 and Table A and Table B of IP 250 referred to in API-ASTM-IP Petroleum Measurement Tables ( API 2540; ASTM D 1250/IP 200 and IP 250 ) has been felt necessary for the purpose of corrections for thermal expansion of the pyknometer in the calculation of density from measurements made at a temperature at which the pyknometer has been calibrated ( Clauses 9.3 and 17.3 ). The Tables mentioned above have, however, not been included in the standard as the same are widely known and in use since long by the various testing laboratories.

In reporting the results of a test or analysis made in accordance with this standard, if the final value, observed or calculated, is to be rounded off, it shall be done in accordance with IS 2 : 1960 'Rules for rounding off numerical values ( revised )'.

*Indian Standard***PETROLEUM AND ITS PRODUCTS —  
METHODS OF TEST**

[ P : 32 ]

**DENSITY AND RELATIVE DENSITY***( Second Revision )***1 SCOPE**

**1.1** Four methods are described to determine the density and relative density of solid and liquid petroleum products and similar materials.

**1.1.1** The application of these methods is indicated in the following table:

reference temperature  $t_2$  is 4°C, relative density and density are numerically equal since the density of water at 4°C is unity.

**3 METHOD A ( RELATIVE DENSITY  
BALANCE METHOD )****3.1 Outline of the Method**

A plummet suspended from the graduated

<i>Product</i>	<i>Recommended Methods</i>
a) Mobile liquids with Reid vapour pressure up to 1.3 kg/cm <sup>2</sup> , e. g., Stabilized crude oil Stabilized Gasoline Naphthas Kerosine Gas oils Lubricating oils Non-waxy fuel oils	A. Density Balance method B. Bicapillary pyknometer method C. Capillary stoppered pyknometer method
b) Viscous liquids and semi-solids, e. g., Waxy fuel oils Slack waxes Soft bitumen	C. Capillary stoppered pyknometer method
c) Solids, e. g., Bitumen	C. (i) Capillary stoppered pyknometer method D. (ii) Displacement method

**2 TERMINOLOGY****2.1 Density of Solids and Liquids**

The density  $\rho$  of a solid or liquid substance at a temperature  $t$  is defined as the mass of the substance occupying unit volume at a temperature  $t$ . When reporting the density the units of mass and volume used together with the temperature shall be explicitly stated, for example, grams per millilitre at  $t^\circ\text{C}$ .

**2.2 Relative Density of Solids and Liquids**

Relative density  $t_1/t_2$  of solid and liquid substance is defined as the ratio of the density of the substance at a temperature  $t_1$  to the density of water at a temperature  $t_2$ . When the

beam of a density or relative density balance ( Fig. 1 ), such as the Westphal balance, is immersed in the liquid which is held at the test temperature. Four rider weights are then disposed on the beam in such a way as to return the latter to its original position. The positions of the riders on the beam gives a direct indication of the relative density of the liquid. The density of the material is calculated from its relative density. The amount of sample required is small ( 15 ml ), and practically all of it is recoverable.

**3.1.1** This method is suitable for determining quickly and accurately the density or relative

density of petroleum products, such as gasoline, kerosine, and gas oil, particularly when the quantities available for test are small.

### 3.2 Apparatus

**3.2.1 Balance** — Westphal or equivalent balance, capable of determining relative density or density to the fourth decimal place, and equipped with a plummet for use at the test temperature.

**3.2.2 Sample Container** — A jacketed container in which the sample temperature, may be kept constant at the test temperature, and from which the sample may be run off at the conclusion of the test.

**3.2.3 Balance Case** — To protect the balance from draughts and dust.

**3.2.4 Support for the Sample Container** — A device for supporting the sample container in a vertical position within the balance case.

#### 3.2.5 Water-Circulating Pump

**3.2.6 Constant-Temperature Bath** — A water-bath capable of maintaining the sample at the test temperature to within  $\pm 0.05^\circ\text{C}$ .

**3.2.7 Thermometer** — See 3.4.1.1.

### 3.3 Buoyancy Corrections

**3.3.1** To obtain the true relative density, that is the relative density in vacuum, of a substance from its relative density in air as determined by comparing the mass in air of equal volumes of the substance and water, a correction shall be applied for buoyancy of the air. This correction for the temperature of  $15^\circ\text{C}$  is given by the following formula:

$$\begin{aligned} \text{The true relative density, } B \\ = A - 0.0012 (A - 1) \end{aligned}$$

where

$$A = \text{relative density in air.}$$

For ready reference, buoyancy corrections for various values of  $[0.0012 (A - 1)]$  are given in Table 1.

### 3.4 Temperature Corrections for Liquid Substances

**3.4.1** Whenever possible the substance whose density or relative density is being determined should be brought to the reporting temperature. When this is not practicable the deviation from the reporting temperature should not exceed  $10^\circ\text{C}$  for obtaining good results. The conversion of density and relative density from the test temperature to  $15^\circ\text{C}$  or

$20^\circ\text{C}$  may be done by using the ASTM/IP Petroleum Measurement Tables.

**3.4.1.1** For the measurement of temperature, suitable thermometer graduated in tenths of a degree shall be used.

### 3.5 General Precautions

**3.5.1** All apparatus shall be absolutely clean and dry before use. When drying and filtering the liquid samples for the removal of water and suspended impurities, precautions shall be taken to minimize loss of volatile constituents.

**3.5.2** When the products are heated to facilitate pouring, changes due to evaporation of volatile constituents shall be avoided.

### 3.6 Standardization

**3.6.1** Assemble the apparatus as shown in Fig. 1 and, using a suitable pump, circulate water from the bath through the annulus of the sample container (1 in Fig. 1) so as to keep the temperature, as measured by the thermometer (2 in Fig. 1) in the container, at the test temperature  $\pm 0.05^\circ\text{C}$ .

**3.6.2** Standardize the instrument in the following way:

Adjust the balance to equilibrium in air, and then fill the container with freshly boiled distilled water. When the latter has reached test temperature (10 to 15 minutes) adjust the riders of the beam (3 in Fig. 1) until equilibrium is obtained and read the result to the fourth decimal place. All subsequent results obtained from the balance shall be divided by this figure.

**3.6.3** Empty the container by opening the tap (4 in Fig. 1) at the bottom; clean it with acetone, drain it, and dry it with a current of air.

### 3.7 Procedure

**3.7.1** Pour the sample to be tested into the clean sample container without splashing, so as to avoid the formation of air bubbles and to reduce to a minimum the evaporation of the lower-boiling constituents of the lighter oils. In the case of reference determinations, errors arising from changes in the volatility may be avoided by transferring the sample to the sample container by water displacement or by siphoning. If air bubbles are formed, remove them after they have collected on the surface by touching them with a piece of clean blotting paper or filter paper before placing the plummet in the sample.




FIG. 1 RELATIVE DENSITY BALANCE

**3.7.2** Make the determination on the material under test using the technique described in 3.6.2 and divide the result by the value obtained for distilled water. This gives the relative density in air of the product at the test temperature.

### 3.8 Calculation and Reporting

**3.8.1** To obtain relative density in vacuo, correct the relative density  $15^{\circ}/15^{\circ}\text{C}$  in air by applying the buoyancy correction from Table 1. Report the value to the fourth decimal place as relative density  $15^{\circ}/15^{\circ}\text{C}$ .

**3.8.2** To obtain the value of density in vacuo, for example at  $15^{\circ}\text{C}$  correct the relative density  $15^{\circ}/15^{\circ}\text{C}$  by applying the buoyancy correction from Table 1, then multiply the figure so obtained by the density of water at  $15^{\circ}\text{C}$  which is  $0.999\text{ g/ml}$ . Report the result to the fourth decimal place.

### 3.9 Precision

Results of duplicate tests shall not differ by more than the following amounts:

Size of Plument	Repeatability	Reproducibility
5 ml	0.000 4	0.000 4

## 4 METHOD B ( GRADUATED BICAPILLARY PYKNOMETER METHOD )

### 4.1 General

**4.1.1** This method covers the determination of the density of crude petroleum and of petroleum products or mixtures of petroleum and non-petroleum products normally handled as liquids. The method is restricted to liquids having Reid vapour pressures of  $1.25\text{ bar}$  ( $1.25 \times 10^5\text{ Pa}$ ) or less and having viscosities less than  $50$  centistokes at the test temperature.



Table 1 Buoyancy Corrections

( Clause 3.3.1 )

[ Corrections to be added are shown with ( + ) mark and those to be subtracted with ( - ) mark ]										
Relative Density at 15/15°C	0.00	0.01	0.02	0.03	0.04	0.05	0.06	0.07	0.08	0.09
0.40	+0.0007	+0.0007	+0.0007	+0.0007	+0.0007	+0.0007	+0.0006	+0.0006	+0.0006	+0.0006
0.50	+0.0006	+0.0006	+0.0006	+0.0006	+0.0006	+0.0005	+0.0005	+0.0005	+0.0005	+0.0005
0.60	+0.0005	+0.0005	+0.0005	+0.0004	+0.0004	+0.0004	+0.0004	+0.0004	+0.0004	+0.0004
0.70	+0.0004	+0.0003	+0.0003	+0.0003	+0.0003	+0.0003	+0.0003	+0.0003	+0.0003	+0.0003
0.80	+0.0002	+0.0002	+0.0002	+0.0002	+0.0002	+0.0002	+0.0002	+0.0002	+0.0002	+0.0001
0.90	+0.0001	+0.0001	+0.0001	+0.0001	+0.0001	+0.0001	0.0000	0.0000	0.0000	0.0000
1.00	0.0000	0.0000	0.0000	0.0000	0.0000	-0.0001	-0.0001	-0.0001	-0.0001	-0.0001
1.10	-0.0001	-0.0001	-0.0001	-0.0001	-0.0002	-0.0002	-0.0002	-0.0002	-0.0002	-0.0002

4.1.2 Special precautions are described for determining the density of highly volatile liquids.

4.1.3 The method is recommended for the accurate determination of the density of all except the more viscous products, and is particularly useful when only small amounts of sample are available.

## 4.2 Outline of the Method


The liquid sample is drawn into the pyknometer and after it has reached equilibrium at the test temperature the liquid levels are noted and the pyknometer is weighed. The density of the sample is then calculated from its weight, a calibration factor proportional to an equal volume of water and a term which corrects for the buoyancy of air.

## 5 APPARATUS

5.1 Pyknometer — capacity 2 to 10 ml conforming to the dimensions given in Fig. 2 constructed of borosilicate or soda-lime glass, annealed after manufacture, and having a total weight not exceeding 30 g. Any bicapillary pyknometer conforming to these requirements may be used.

5.2 Constant Temperature Bath — water bath having a depth of at least 300 mm, capable of being maintained within  $\pm 0.05^{\circ}\text{C}$  of the desired temperature.

5.3 Bath Thermometer — See 3.4.1.1. Other total immersion thermometers of suitable range and equal or greater accuracy may also be used.



All dimensions in millimetres.

FIG. 2 GRADUATED BICAPILLARY PYKNOMETER

**5.4 Pyknometer Holder (Optional)** — to allow the pyknometer to be mounted vertically at the correct depth in the constant temperature bath. A suitable design is shown in Fig. 3; it shall be constructed of metal which will not corrode in the water bath. A series of pyknometer holders may be supported in the constant-temperature bath by the use of a non-corrodible rectangular metal bar of sufficient length to lie across the rim of the bath. A series of holes of sufficient diameter to accommodate the 6.5 mm rod of the pyknometer holder is drilled in the bar at about 45 mm apart. Each rod is secured in its hole by locking the bar between the hexagon nut and the wing nut and washer.

Alternatively any other form of pyknometer mounting may be employed which allows the pyknometer to be held vertically in the constant temperature bath at the correct depth of immersion.

**5.5 Balance** — capable of weighing to the nearest 0.1 mg.

## 6 PREPARATION OF APPARATUS

**6.1** Thoroughly clean the pyknometer with surfactant cleaning fluid, rinse well with distilled water, then with acetone and dry. Ensure that all traces of moisture are removed by drying with a current of filtered air passing slowly through the pyknometer. Wipe the outside of the pyknometer with a clean, lint-free cloth. Cleaning should be carried out in this manner whenever the pyknometer is to be calibrated or whenever liquid fails to drain cleanly from the internal walls of the pyknometer. Normally the pyknometer may be cleaned between determinations by washing with a suitable light petroleum solvent, followed by vacuum drying.

**NOTE** — If sufficient cleaning fluids do not give adequate cleaning, chromic acid cleaning solution may be used.


**CAUTION:** Chromic acid is potentially hazardous in contact with organic materials and is toxic and highly corrosive. If used wear approved face-shield and full-length protecting clothing including gloves.

## 7 CALIBRATION OF APPARATUS

**7.1** After drying allow the pyknometer to reach room temperature. Dissipate any static charge which may have formed and weigh to the nearest 0.1 mg.

**NOTE** — If the balance is not fitted with a static eliminator, static charges may be dissipated by breathing on the pyknometer. In such cases ensure that the pyknometer has regained constant mass before recording the weight.

**7.2** Fill the pyknometer with sufficient freshly boiled, cooled, distilled water to obtain readings near the top of the graduated capillaries. Filling is readily achieved by placing the hooked tip in the liquid while keeping the pyknometer upright, thus allowing the liquid to be drawn over the bend in the capillary attraction.



All dimensions in millimetres.

**FIG. 3** A SUITABLE DESIGN OF PYKNOMETER HOLDER

The pycnometer then fills by syphoning. Place the pycnometer in the constant temperature bath so that the whole of the liquid in the pycnometer is below the level of the bath liquid. Maintain the temperature of the bath at  $15^{\circ} \pm 0.05^{\circ}\text{C}$  or within  $\pm 0.05^{\circ}\text{C}$  of the required calibration temperature. Keep the pycnometer in the bath for 20 min then read the scale to the nearest small division at the liquid level in each arm.

**7.3** Remove the pycnometer from the bath, allow the water on the exterior to drain off and wipe with a clean, dry, lint-free cloth. Allow to come to room temperature, dissipate any static charge, and weigh to the nearest 0.1 mg.

**7.4** The difference between the weights of the filled and empty pycnometer gives the weight of water contained by the pycnometer at the test temperature, corresponding to the some of the two scale readings. By removing successive quantities of water repeat the determination so as to obtain a series of at least three pairs of readings, together with the corresponding weights, with the water level at different scale points on the graduated arms. One pair of readings should be at the upper end of the scale and another at the lower end. Plot the sums of the scale readings on the two arms against the corresponding weights. The points should lie on a straight line which gives the weight of water contained by the pycnometer for any combined scale reading. If the points show a scatter of more than two small scale divisions as marked on the pycnometer arms on either side of a straight line drawn through the array of points and subsequent test do not correct this, discard the pycnometer as imperfect.

**7.5** If it is desired to determine density at a temperature other than  $15^{\circ}\text{C}$  calibrate the pycnometer at the desired temperature.

**7.6** Re-calibrate pycnometer at intervals as dictated by experience.

NOTE — It is recommended that pycnometer be calibrated at the test temperature at which densities are to be determined, since this greatly simplifies the calculation of densities and eliminates any uncertainty as to the coefficient of cubical expansion of the glass used in the manufacture of the pycnometer.

## 8 PROCEDURE

**8.1** Weigh the clean, dry pycnometer to the nearest 0.1 mg, dissipating any static charge if necessary ( see Note under 7.1 ).

**8.2** Fill the pycnometer with the sample at

approximately the test temperature by the method described in 7.2, so that the liquid levels are in the graduated portions of the capillaries. When the test temperature is lower than the laboratory temperature low scale readings should be aimed at in order to minimize any losses due to evaporation during weighing. Bring the pycnometer and contents to the test temperature  $t_1$  by immersion for 20 minutes in the constant temperature both as described in 7.2 and obtain readings of the liquid level in the two graduated arms. In the case of the more viscous and/or opaque samples no readings must be taken until ample time for draining has been allowed after any disturbance of the pycnometer. The 20 minutes immersion time is normally sufficient, provided that the pycnometer has not been disturbed during this period.

NOTE — For mixtures of non-petroleum products it is essential to ensure that the test temperature is the same as the final reporting temperature unless an approximate value is acceptable and the volumetric compression of the mixture is known together with the volumetric correction coefficient of the components in the mixture.

**8.3** Remove the pycnometer from the bath, allow the water on the exterior to drain off and wipe with a clean, dry, lint-free cloth. Allow to come to room temperature, dissipate any static charge, and weigh to the nearest 0.1 mg.

**8.4** When carrying out the determination on highly volatile samples containing appreciable amounts of material boiling below  $20^{\circ}\text{C}$ , or for any sample where there is uncertainty concerning loss which might result from evaporation during the determination, cool the sample and pycnometer to a temperature of  $0^{\circ}$  to  $5^{\circ}\text{C}$  before filling. If the dew point is sufficiently high to cause condensation of moisture in the pycnometer during the cooling operation, attach a drying tube to the arm of the pycnometer in order to avoid this. With samples of this type it is essential to restrict the filling of the pycnometer to obtain a low scale reading thus minimizing losses due to evaporation. If the total length of unfilled capillary is over 10 cm the rate of diffusion is so low that even with highly volatile compounds such as *iso*-pentane vapour losses during the determination are negligibly low.

## 9 CALCULATION AND REPORTING

### 9.1 Symbols

The following symbols are used in the calculations:

$t_a$  — any reference temperature, e.g.  $15^{\circ}\text{C}$ ;

$t_c$  — the temperature at which the pycnometer is calibrated by water filling ( see 7.2 and 7.5 );

$t_t$  — the temperature at which the pycnometer is filled with the liquid under test ( see 9.2.2 );

$m_o$  — the apparent mass in air ( weight in air ) ( g ) of the empty pycnometer;

$m_c$  — the apparent mass in air ( g ) of the pycnometer filled with water at the calibration temperature  $t_c$  ;

$m_t$  — the apparent mass in air ( g ) of the pycnometer filled with the liquid under test at the temperature  $t_t$  ;

$C$  — the correction for air buoyancy, in g/ml ( see Table 1 );

$\rho_c$  — the density of water, in kg/m<sup>3</sup>, at the temperature of calibration  $t_c$  ( see Table 2 );

$\alpha_1$  — the coefficient of cubical expansion of borosilicate glass ( see 9.3.2 );

$\alpha_2$  — the coefficient of cubical expansion of soda-lime glass ( see 9.3.3 );

$\rho_t$  — the density of the sample, in g/ml, at the test temperature  $t_t$  ;

$\rho_s$  — the density of the sample in g/ml, at any reference temperature  $t_s$  ;

$\rho_{15}$  — the density of the sample, in g/ml, at the reference temperature of 15°C;

$\rho_{20}$  — the density of the sample, in g/ml, at the reference temperature of 20°C;

$\rho_t$  — the apparent density, in g/ml, at the test temperature  $t_t$  as determined in soda-lime glass apparatus calibrated at the reference temperature  $t_s = 15^\circ$  or  $20^\circ\text{C}$ , i.e. the observed density uncorrected for glass expansion — required for entering the API-ASTM — IP Petroleum Measurement Tables.

NOTE — The conventional unit g/ml has been retained for this calibration although the API-ASTM-IP Petroleum Measurement Tables for density are given in kg/cm<sup>3</sup>.

## 9.2 Reference, Calibration and Test Temperatures

9.2.1 The pycnometer may be calibrated at any convenient temperature and this may correspond with the reference or test temperatures. It is recommended that pycnometers be calibrated at the test temperature at which densities are to be determined since this greatly simplifies the calculation of density and only the calculations given in 9.4.1.1, 9.4.2.1 and 9.4.2.2 then apply.

Table 2 Buoyancy Corrections\*

( Clause 9.1 )

$\frac{m_t - m_o}{m_c - m_o}$	Correction, g/ml	$\frac{m_t - m_o}{m_c - m_o}$	Correction g/ml
0.60	0.000 48	0.80	0.000 24
0.61	0.000 47	0.81	0.000 23
0.62	0.000 46	0.82	0.000 22
0.63	0.000 44	0.83	0.000 20
0.64	0.000 43	0.84	0.000 19
0.65	0.000 42	0.85	0.000 18
0.66	0.000 41	0.86	0.000 17
0.67	0.000 40	0.87	0.000 16
0.68	0.000 38	0.88	0.000 14
0.69	0.000 37	0.89	0.000 13
0.70	0.000 36	0.90	0.000 12
0.71	0.000 35	0.91	0.000 11
0.72	0.000 34	0.92	0.000 10
0.73	0.000 32	0.93	0.000 08
0.74	0.000 31	0.94	0.000 07
0.75	0.000 30	0.95	0.000 06
0.76	0.000 29	0.96	0.000 05
0.77	0.000 28	0.97	0.000 04
0.78	0.000 26	0.98	0.000 02
0.79	0.000 25	0.99	0.000 01

\*Calculated for standard air of density 0.001 222 g/ml at 15.56°C and at a pressure of 1.013 bar.

These corrections are applicable for air density values between 0.001 1 and 0.001 3 g/ml.

## 9.3 Corrections for the Thermal Expansion of the Pycnometer

### 9.3.1 General

The calculation of density from measurements made at the temperature  $t_t$  which differs from the temperature  $t_c$  at which the pycnometer was calibrated, involves a correction for cubical expansion or contraction of the glass from which the pycnometer is made.

If the calculation is based on the density correction tables given in the API - ASTM - IP Petroleum Measurement Tables or IP 250, a similar correction may also be required ( see 9.3.4 ).

### 9.3.2 Pycnometer Made of Borosilicate Glass

9.3.2.1 The coefficient of cubical expansion of borosilicate glasses are known to depend on the glass to fall three main categories having coefficient of cubical expansion of  $10 \times 10^{-6}$ ,  $14 \times 10^{-6}$  and  $19 \times 10^{-6}$  per 1°C respectively.

For determination of the highest accuracy when borosilicate pyknometers are used, therefore, either

- a) ensure that  $t_t = t_c'$  or
- b) use a pyknometer for which the coefficient of cubical expansion is known. When the foregoing is not possible and a lower precision is acceptable, the use of  $10 \times 10^{-6}$  per  $1^\circ\text{C}$  is recommended.

### 9.3.3 Pyknometers, Made of Soda-lime Glass

9.3.3.1 For pyknometers made of soda-lime glass, the coefficient of cubical expansion may be assumed to be  $25 \times 10^{-6}$  per  $1^\circ\text{C}$ .

### 9.3.4 API-ASTM-IP Petroleum Measurement Tables (ASTM D 1250 IP and IP 250)

9.3.4.1 The following tables referred to in API-ASTM-IP Petroleum Measurement Tables are relevant only for determinations of petroleum and petroleum products. The tables must not be used for non-petroleum products for which it must always be ensured that  $t_t = t_s$ ,  $\rho_s$  is required.

IP 200 Table 53 Correction of Observed Density to Density at  $15^\circ\text{C}$ .

IP 250 Table A Correction of Observed Density to Density at  $20^\circ\text{C}$ .

IP 200 Table 54 Correction of Observed Volume to Volume at a reference temperature of  $15^\circ\text{C}$ .

IP 250 Table B Correction of Observed Volume to Volume at a reference temperature of  $20^\circ\text{C}$ .

9.3.4.2 Enter Tables 53 A, 53B, 53D or Table A (according to the type of sample under test) with densities determined in glass apparatus calibrated at  $15^\circ$  or  $20^\circ\text{C}$  as appropriate.

The tables include a correction for the expansion of the glass apparatus used (assumed to be soda-lime glass) for which the conventional value of  $23 \times 10^{-6}$  per  $1^\circ\text{C}$  applies.

9.3.4.3 When entering Tables 53A, 53B, 53D or Table A with the results of determinations made with pyknometers constructed from borosilicate glass, a correction will be required for the difference between the coefficient of cubical expansion of the borosilicate glass and the conventional value for soda-lime glass incorporated in the tables.

9.3.4.4 When the pyknometer used is made of soda-lime glass no correction is necessary when

using Tables 53A, 53B, 53D or Table A when  $t_c = 15^\circ$  or  $20^\circ\text{C}$  as appropriate to the table entered.

If  $t_c$  does not correspond to the reference temperature  $t_s$  for which the table used is constructed, an adjustment is required for the expansion of the pyknometer over the temperature range  $t_c \rightarrow t_s$ .

9.3.4.5 Alternatively, the corresponding volume correction Tables 54A, 54B, or 54D may be used. These tables do not include any correction for the expansion of glass but must be entered with a density at the appropriate reference temperature. Consequently the use of these tables requires a tedious repetitive calculation. This procedure is not therefore recommended and is not covered in 9.4 and 9.5.

## 9.4 Calculation of Density

### 9.4.1 Density at any Temperature $t_t$

9.4.1.1 When  $t_t = t_c$ ,

$$\rho_t = \left[ \frac{(m_t - m_o)}{(m_c - m_o)} 1000 \right] + C$$

9.4.1.2 When  $t_t = t_c$ ,

$$\rho_t = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000 [1 - \alpha^1 (t_c - t_t)]} \right] + C$$

where  $\alpha^1 = \alpha_1$  or  $\alpha_2$  as appropriate to the pyknometer used.

### 9.4.2 Density at Reference Temperature, $t_s$

9.4.2.1 When  $t_t = t_c = t_s$ ,

$$\rho_s = \rho_t = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C$$

9.4.2.2 When  $t_t = t_c \neq t_s$  ( $t_s = 15^\circ$  or  $20^\circ\text{C}$ )

The density determined at the test temperature  $t_t$  as in 9.4.1.1 is the true density at  $t_t$  in soda-lime glass apparatus calibrated at  $t_s$  before entering Tables 53A, 53B, 53D or Table A as appropriate. This always applies irrespective of the glass from which the pyknometer is made, i.e. enter Tables 53A, 53B, 53D or Table A with  $\rho_t^1$  and against  $t_t$  extract  $\rho_{15}$  or  $\rho_{20}$  respectively, interpolating as necessary, where

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [1 + \alpha_2 (t_s - t_t)]$$

**9.4.2.3** When  $t_c = t_s \neq t_t$  ( $t_s = 15^\circ$  or  $20^\circ\text{C}$ )

- a) Pyknometer made of soda-lime glass — Tables 53A, 53B, 53D or Table A may be entered directly with  $\rho_t^1$  and against  $t_t$  the corresponding  $\rho_{15}$  or  $\rho_{20}$  is extracted, interpolating as necessary, where

$$\rho_t^1 = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C$$

- b) Pyknometer made of borosilicate glass — Tables 53A, 53B, 53D or Table A must be entered with  $\rho_t^1$  adjusted for the differential expansion of the pyknometer glass and against  $t_t$  the corresponding  $\rho_{15}$  or  $\rho_{20}$  is extracted

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [ 1 + (\alpha_2 - \alpha_1) (t_s - t_t) ]$$

Second-order terms, which are not significant, are ignored.

**9.4.2.4** When  $t_t \neq t_c \neq t_s$ 

The density determined at the test temperature  $t_t$  as in 9.4.1.2 is the true density at  $t_t$ . The result obtained must, therefore, be adjusted to give the apparent density at  $t_t$  in soda-lime glass apparatus calibrated at  $t_s$  before entering Tables 53A, 53B, 53D or Table A as appropriate. This always applies, irrespective of the glass from which the pyknometer is made, i.e. enter Tables 53A, 53B, 53D or Table A with  $\rho_t^1$  and against  $t_t$  extract  $\rho_{15}$  or  $\rho_{20}$  as appropriate, interpolating as necessary. The expression for  $\rho_t^1$  may be rearranged as follows, ignoring second-order terms which are not significant.

- a) Pyknometers made of soda-lime glass —

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [ 1 + \alpha_2 (t_s - t_c) ]$$

- b) Pyknometers made of borosilicate glass —

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [ 1 + \alpha_1 t_s + \alpha_2 t_t - (\alpha_1 + \alpha_2) t_c ]$$

**9.4.2.5** When  $t_s = t_t \neq t_c$ 

Use the expression in 9.4.1.2.

**9.4.3 Rounding-Off of Calculations and Reporting**

Carry out all calculations to five figures and report the final result to the nearest 0.0001 as density giving the units of mass and volume and the temperature.

**10 PRECISION**

**10.1** The following criteria should be used for judging the acceptability of results (95 percent confidence).

**10.1.1 Repeatability**

Duplicate results by the same operator should be considered suspect if they differ by more than the following amounts:

Range, g/ml	Repeatability
0.777 0 to 0.892 0	0.000 7

**10.1.2 Reproducibility**

The results obtained by each of two laboratories should not be considered suspect unless they differ by more than the following amounts:

Range, g/ml	Reproducibility
0.777 0 to 0.892 0	0.001 0

The above values were obtained when using 5 ml pyknometers. The use of larger (e.g. 10 ml) pyknometers should give equal or better accuracy.

The method is not restricted to samples having densities within the range 0.777-0.892 g/ml but precision figures are not available for values outside this range.

**11 METHOD C (CAPILLARY-STOPPERED PYKNOMETER METHOD)****11.1 General**

**11.1.1** This method covers the determination of the density or relative density of crude petroleum and of petroleum products handled as liquids or solids. The method may also be used for coal tar products, including road tars, creosote and tar pitches, or for mixtures of these with petroleum products.

**11.1.2** The method is not suitable for the determination of the density or relative density of highly volatile liquids having Reid vapour pressure greater than 0.5 bar ( $5 \times 10^4$  Pa) or having initial boiling points below  $40^\circ\text{C}$ .

**11.2 Summary of Method**

**11.2.1** The weights of equal volumes of the sample and of water are compared. Equal volumes are ensured by placing the filled pyknometer in a bath at the test temperature until equilibrium is reached.

## 12 APPARATUS

**12.1 Pyknometer** — One of the four types shown in Fig. 4 and 5 ( *see* 15.1 and 16.1 ), constructed of borosilicate glass or soda-lime glass and annealed after manufacture.

The 'warden' form [ Fig. 4 ( b ) ] is determined for all except viscous or solid products and should always be used for volatile products. The ground-glass cap, or 'warden', almost eliminates expansion and evaporation losses and this form of pyknometer may be used when the test temperature is lower than that of the laboratory.

The conical ( Fig. 5 ) is recommended for waxy oils and semi-solids.

The form of pyknometer shown in Fig. 4 ( a ) is suitable for non-volatile liquids except those of high viscosity.

The wide-mouth ( Hubbard ) form of pyknometer [ Fig. 4 ( c ) ] is used for very viscous liquids and solids.

As the forms of pyknometer shown in Fig. 4 ( a ) and ( c ) have no 'warden' or expansion chamber, they cannot be used when the temperature of the test is so far below that of the laboratory as to cause loss of sample by expansion through the capillary.

**12.2 Constant-temperature Bath** — a water bath having a depth greater than that of the pyknometer, capable of being maintained within  $\pm 0.05^{\circ}\text{C}$  ( $0.1^{\circ}\text{F}$ ) of the desired temperature.

**12.3 Bath Thermometer** — *See* 3.4.1.1. Other total immersion thermometers of suitable range and equal or greater accuracy may also be used.

**12.4 Pyknometer Holder ( Optional )** — to hold pyknometer vertically and at the correct depth in the constant-temperature bath, a metal holder is desirable. It may be constructed of any suitable metal which will not corrode in the water bath.

**12.5 Balance** — capable of weighing to the nearest 0.1 mg.

## 13 PREPARATION OF APPARATUS

**13.1** Thoroughly clean the pyknometer and stopper with a surfactant cleaning fluid, rinse well with distilled water, then with acetone and dry. Ensure that all traces of moisture are removed by drying with a current of filtered air passing slowly through the pyknometer and stopper capillary. Wipe the outside of the pyknometer and stopper with a clean, lint-free cloth.




FIG. 4 PYKNOMETER FORMS




FIG. 5 CONICAL CAPILLARY-STOPPER PYKNOMETER

**NOTE** — If sufficient cleaning fluids do not give adequate cleaning, chromic acid cleaning solution may be used.

**CAUTION** — Chromic acid is potentially hazardous in contact with organic materials and is toxic and highly corrosive. If used, wear approved full face-shield and full-length protective clothing including gloves.

Cleaning should be carried out in this manner whenever the pyknometer is to be calibrated or whenever liquid fails to drain cleanly from the internal walls of the pyknometer or the capillary of the stopper. Normally the pyknometer may be cleaned between determinations by washing with a suitable light petroleum solvent, followed by vacuum drying.

#### 14 CALIBRATION OF PYKNOMETER

**14.1** After drying allow the pyknometer to reach room temperature. Dissipate any static charge which may have formed on it and then weigh, complete with warden if of form 4 (b) or form 5, to the nearest 0.1 mg.

**NOTE** — If the balance is not fitted with a static eliminator, static charges may be dissipated by breathing on the pyknometer. In such cases ensure that the pyknometer has regained constant mass before recording the weight.

**14.2** Fill the pyknometer with freshly-boiled

distilled water, cooled to slightly below 15°C, and firmly insert the stopper, taking care to avoid the inclusion of any air bubbles. Immerse the pyknometer to the neck in the constant-temperature bath. When using the conical form of pyknometer, place the reservoir cap in the constant-temperature bath. Maintain the bath within  $\pm 0.05^\circ\text{C}$  ( $\pm 0.1^\circ\text{F}$ ) of the calibration temperature for not less than one hour.

**14.3** With the pyknometer and its contents at this temperature wipe the top of the stopper so that it is dry and the meniscus of the water in the capillary is flush with the top of the stopper. Care is necessary during this operation, since capillary action of the cloth can draw liquid out of the stopper. Place the warden firmly on the stopper (if the pyknometer is of this type). When using the conical form, raise the pyknometer in the water bath until the smaller cone is clear of the water. Remove the reservoir cap and wipe the cone so that it is dry and the meniscus of the water in the capillary is flush with the top of the stopper, then place the warden firmly on the stopper.

**14.4** Remove the pyknometer from the bath. If not of the 'warden' form, cool the pyknometer and its contents to a temperature slightly below the temperature of the bath.



**14.5** Dry the exterior surface of the pyknometer by wiping with a clean, lint-free cloth, paying particular attention to the joint between the stopper and pyknometer, dissipate any static charge and weigh to the nearest 0.1 mg.

**14.6** The difference between the weights of the filled and empty pyknometer gives the mass of water contained by the pyknometer at the calibration temperature.

**14.7** If it is required to determine density at a temperature other than 15°C or to determine relative density referred to water at a temperature other than (60°F) 15.6°C, calibrate the pyknometer at the required temperature. It is recommended that pyknometer be calibrated at the test temperature at which the densities are to be determined, since this greatly simplifies the calculation of densities and relative densities and eliminates any uncertainty as to the type of glass used in the construction of the pyknometer.

**14.8** Re-calibrate pyknometers at intervals as dictated by experience.

NOTE — It is recommended that new pyknometers should be re-calibrated after one year and thereafter at intervals dependent upon the magnitude of any changes found. Soda-lime glass pyknometers may require calibration more frequently than borosilicate glass pyknometers.

## 15 PROCEDURE FOR LIQUIDS

**15.1** Choose an appropriate form and size of pyknometer for the sample to be tested (see 12.1). The 75 ml and 50 ml sizes are normally the most suitable.

**15.2** Weigh the clean, dry pyknometer with warden [if of form 4 (b) or 5] and if necessary, disperse any static charge (see Note given under 14.1). Pyknometers of 25 ml or greater capacity should be weighed to the nearest 0.5 mg, and those of smaller capacity to the nearest 0.1 mg.

**15.3** Fill the pyknometer with the sample, if necessary warming both sample and pyknometer to assist filling and separation of air bubbles. Bring the pyknometer and its contents to the test temperature ( $t_t$ ) by immersing pyknometer up to its neck in the constant-temperature bath. Keep the pyknometer in the bath for 20 minutes in order to stabilize the temperature and to permit air bubbles to rise to the surface. If after this time the liquid level is still changing, keep the pyknometer in the bath until the liquid level becomes stable.

NOTE — For mixtures containing non-petroleum products it is essential to ensure that the test temperature is the same as the final reporting temperature unless an appropriate value is acceptable and the volumetric composition of the mixture is known, together with the volumetric correction coefficients of the components in the mixture.

When the temperature is constant, firmly insert the dry capillary stopper, which has also been brought to the test temperature, taking care to avoid trapping air bubbles below the stopper.

NOTE — It is essential to ensure that no air bubbles are left trapped in the liquid and adequate time must be allowed for air bubbles to rise to the surface before inserting the stopper.

Wipe excess liquid from the top of the stopper so that the meniscus of the liquid in the capillary is flush with the top of the stopper. Place the 'warden' over the stopper (if the pyknometer is of this type).

**15.4** Remove the pyknometer from the bath and, if not of the 'warden' type cool to a temperature slightly below  $t_t$ . Cool the pyknometer and contents to room temperature if the test temperature is high.

**15.5** Remove all traces of sample and water from the exterior surface of the pyknometer by wiping with a clean, lint-free cloth, disperse any static charge and weigh to the precision given in 15.2.

## 16 PROCEDURE FOR SOLID OR SEMISOLID SAMPLES

**16.1** Weigh the clean, dry pyknometer which should be of the wide-mouth type [Fig 4 (c)] or the conical type (Fig. 5) complete with warden to the nearest 0.5 mg.

**16.2** Introduce a suitable amount of the sample in the form of small pieces, which should be as regular in order to reduce the possibility of trapping air bubbles. Alternatively, pour the molten sample in to the warmed pyknometer, taking care to avoid the inclusion of air bubbles. Fill the conical pyknometer up to the 20 ml graduation mark. Allow the test portion and pyknometer to cool slowly to ambient temperature. Check that no voids have formed between the test portion sufficiently to remove the voids and again allow to cool.

**16.3** Bring the pyknometer and its contents to room temperature and weigh to the nearest 0.5 mg.

**16.4** In cases of sample having very high wax contents it may be necessary to weigh the cooled test portion and pycnometer before overcoming the problem of voids. The weighed test portion and pycnometer are then reheated until the voids disappear and allowed to cool. As soon as the melted test portion has set, the surface should be covered with freshly boiled and cooled distilled water to prevent the ingress of air into any voids which may form later at the bottom of the pycnometer.

**16.5** Fill the pycnometer with freshly-boiled, cooled distilled water, removing all air bubbles. A fine wire may be used to facilitate the removal of bubbles. Bring the pycnometer and its contents to the test temperature, which must be the same as the calibration temperature of the pycnometer, by immersing the pycnometer up to its neck in the constant-temperature bath. When using the conical form of pycnometer, place the reservoir cap on the capillary stopper and firmly insert it, taking care to avoid trapping air bubbles below the stopper, and immerse to half-way up the reservoir cap in the constant-temperature bath.

Keep the pycnometer in the bath for 20 minutes in order to stabilize the temperature and to permit bubbles to rise to the surface. If after this time the liquid level is still changing, keep the pycnometer in the bath until the liquid level becomes stable.

When the temperature is constant, firmly insert the capillary stopper, which has also been brought to the test temperature, taking care to avoid trapping air bubbles below the stopper. When using the conical pycnometer it is only necessary to raise the pycnometer until the smaller cone is clear of the water and remove the reservoir cap. Wipe excess water from the top of the stopper so that the meniscus of the water in the capillary is flush with the top of the stopper. Place the warden firmly on the stopper if the pycnometer is of the conical form.

**16.6** Remove the pycnometer from the bath and if the pycnometer is of the wide-mouth form, cool to a temperature slightly below the tests/calibration temperature. Cool the pycnometer and contents to room temperature if the test temperature is above ambient.

**16.7** Dry the exterior surface of the pycnometer by wiping with a clean, lint-free cloth, disperse any static charge and weigh to the nearest 0.5 mg.

## 17 CALCULATION AND REPORTING

**17.1 Symbols** — The following symbols are used in the calculations :

- $t_s$  — any reference temperature, e. g. 15°C;
- $t_c$  — the temperature at which the pycnometer is calibrated by water filling ( *see* 16.5 and 17.2 );
- $t_t$  — the temperature at which the pycnometer is filled with liquid under test ( *see* 17.2.2 );
- $m_o$  — the apparent mass in air ( weight in air ) ( g ) of the empty pycnometer;
- $m_c$  — the apparent mass in air ( g ) of the pycnometer filled with water at the calibration temperature  $t_c$ ;
- $m_t$  — the apparent mass in air ( g ) of the pycnometer filled with the liquid under test at the temperature  $t_t$ ;
- $m_1$  — the apparent mass in air ( g ) of the pycnometer plus solid or semi-solid ( *see* 16.3 );
- $m_2$  — the apparent mass in air ( g ) of the pycnometer plus solid or semi-solid plus water ( *see* 16.7 );
- $C$  — the correction of air buoyancy, in g/ml ( *see* Table 4 );
- $\rho_c$  — the density of water, in kg/m<sup>3</sup>, at the temperature of calibration;
- $\alpha_1$  — the coefficient of cubical expansion of borosilicate glass ( *see* 17.3.2 );
- $\alpha_2$  — the coefficient of cubical expansion of soda-lime glass ( *see* 17.3.3 );
- $\rho_t$  — the density of the sample, in g/ml, at the test temperature  $t_t$ ;
- $\rho_s$  — the density of sample in g/ml, at any reference temperature  $t_s$ ;
- $\rho_{15}$  — the density of the sample, in g/ml, at the reference temperature of 15°C;
- $\rho_{20}$  — the density of the sample, in g/ml, at the reference temperature of 20°C;
- $\rho_t^1$  — the apparent density, in g/ml, at the test temperature  $t_t$  as determined in soda-lime glass apparatus calibrated at the reference temperature  $t_s = 15^\circ$  or  $20^\circ\text{C}$ , i. e., the observed density uncorrected for glass expansion — required for entering the API-ASTM-IP Petroleum Measurement Tables.
- $d_t$  — the relevant density at the test temperature  $t_t$ ;
- $d_s$  — the relevant density at reference temperature  $t_s$ ;

$d_{15.6}$  — the relevant density at the reference temperature of ( 60°F ) 15.6°C;

$d_t^1$  — the apparent relative density at the test temperature  $t_t$  as determined in glass apparatus calibrated at the reference temperature  $t_s = ( 60°F ) 15.6°C$ , i. e. the observed relative density uncorrected for glass expansion required for entering the API-ASTM-IP Petroleum Measurement Tables.

## 17.2 Reference, Calibration and Test Temperatures

**17.2.1** The pycnometer may be calibrated at any convenient temperature and this may correspond with the reference or test temperatures. The calculation of density is greatly simplified when the pycnometer is calibrated at the test temperature since only the calculations given in 17.4.1.1, 17.4.2.1 and 17.4.2.2 then apply.

**17.2.2** For qualitative purposes, the test temperature is usually chosen to correspond with the required reference temperature, but for quantitative purposes involving the calculation of the mass or of the apparent mass in air of a given quantity of oil, the density or relative density should be determined within 3°C of the temperature at which the oil is measured by the selected dynamic or static method. However, in the case of very volatile products having a Reid vapour pressure above 0.5 bar, the test should be carried out at a temperature of 15°C or below to minimize the loss of light fractions.

## 17.3 Correction for the Thermal Expansion of the Pycnometer

### 17.3.1 General

The calculation of density or of relative density from measurements made at a temperature  $t_t$  which differs from the temperature  $t_c$  at which the pycnometer was calibrated, involves a correction for cubical expansion or contraction of the glass from which the pycnometer is made.

If the calculation is based on the density or relative density correction tables given in the API-ASTM-IP Petroleum Measurement Tables, a similar correction may also be required ( see 17.3.4 ).

### 17.3.2 Pycnometers Made of Borosilicate Glass

**17.3.2.1** The coefficient of cubical expansion of borosilicate glass are known to depend on the

source of the glass and to fall into three main categories having coefficients of cubical expansion of  $10 \times 10^{-6}$ ,  $14 \times 10^{-6}$  and  $19 \times 10^{-6}$  per 1°C respectively.

For determination of the highest accuracy when borosilicate pycnometers are used, therefore, either

- a) ensure that  $t_t = t_c$  or
- b) use a pycnometer for which the coefficient of cubical expansion is known.

When the foregoing is not possible and a lower precision is acceptable, the use of  $10 \times 10^{-6}$  per 1°C is recommended.

### 17.3.3 Pycnometers Made of Soda-Lime Glass

**17.3.3.1** For pycnometers made of soda-lime glass, the coefficient of cubical expansion may be assumed to be  $25 \times 10^{-6}$  per 1°C.

### 17.3.4 The API-ASTM-IP Petroleum Measurement Tables ( API 2540 : ASTM D 250/IP 200 and IP 250

**17.3.4.1** The following tables referred to in the API-ASTM-IP Petroleum Measurement Tables are relevant only for determinations non-petroleum and petroleum products. The tables must not be used for non-petroleum products for which it must always be ensured that  $t_t = t_s$  when  $\rho_s$  or  $d_s$  is required:

IP 200 Tables 53A, 53B and 53D	Correction of Observed Density to Density at 15°C
IP 250 Table A	Correction of Observed Density to Density at 20°C
IP 200 Tables 23A, 23B and 23D	Correction of observed Relative Density (60°/60°F) 15.6°/15.6°C to Relative Density ( 60°/60°F ) 15.6°/15.6°C
IP 200 Tables 54A, 54B and 54D	Correction of Observed Volume to Volume at a reference temperature of 15°C
IP 250 Table B	Correction of Observed Volume to Volume at a reference temperature of 20°C
IP 200 Tables 24A, 24B and 24D	Correction of Observed Volume to Volume at a reference temperature

**17.3.4.2** Tables 53A, 53B, A, 53D, 23A, 23B and 23D are entered with densities or relative densities determined in glass apparatus calibrated at a temperature  $t_c$  in °C ( Tables 53A, 53B 53D or A ) or °F ( Tables 23A, 23B or 23D ).

The tables include a correction for the expansion of the glass apparatus used ( assumed to be soda-lime glass ) for which the conventional value of  $( 25 \pm 2 ) \times 10^{-6}$  per 1°C applies.

**17.3.4.3** When entering Tables 53A, 53B, 53D, A, 23A, 23B, 23D with the results of determinations made with pycnometers constructed from borosilicate glass, a correction will be required for the difference between the coefficient of cubical expansion of the borosilicate glass and the conventional value for soda-lime glass incorporated in the tables.

**17.3.4.4** When the pycnometer used is made of soda-lime glass no correction is necessary when using Tables 53A, 53B, 53D, A, 23A or 23B, when  $t_c = 15^\circ\text{C}$ ,  $20^\circ\text{C}$  or  $( 60^\circ\text{F} ) 15.6^\circ\text{C}$  as appropriate to the table entered.

If  $t_c$  does not correspond to the reference temperature  $t_s$  for which the table used is constructed, an adjustment is required for the expansion of the pycnometer over the temperature range  $t_c \rightarrow t_s$ .

**17.3.4.5** Alternatively, the corresponding volume correction Tables 54A, 54B, 54D, 24A, 24B or 24D may be used. These tables do not include any correction for the expansion of glass, but must be entered with a density at the appropriate reference temperature. Consequently the use of these tables requires a tedious repetitive calculation. This procedure is not, therefore, recommended and is not covered in 17.4 and 17.5.

## 17.4 Calculation of Density

### 17.4.1 Density at any Temperature $t_c$

#### 17.4.1.1 When $t_t = t_c$

a) for liquids

$$\rho_t = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C$$

b) for solids or semi-solids

$$\rho_t = \frac{(m_1 - m_o) \rho_c}{(m_c - m_o - m_2 + m_1) 1000} + C$$

#### 17.4.1.2 When $t_t \neq t_c$

$$\rho_t = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000 [ 1 - \alpha^1 ( t_c - t_t ) ]} + C$$

where  $\alpha^1 = \alpha_1$  or  $\alpha_2$  as appropriate to the pycnometer used.

### 17.4.2 Density at Reference Temperature $t_s$

#### 17.4.2.1 When $t_t = t_c = t_s$

a) for liquids

$$\rho_s = \rho_t = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C$$

b) for solids or semi-solids

$$\rho_s = \rho_t = \frac{(m_1 - m_o) \rho_c}{(m_c - m_o - m_2 + m_1) 1000} + C$$

#### 17.4.2.2 When $t_t = t_c \neq t_s = 15^\circ$ or $20^\circ\text{C}$

The density determined at the test temperature  $t_t$  as in 17.4.1.1, is the true density at  $t_t$ . The result obtained must, therefore, be adjusted to give the apparent density at  $t_t$  in soda-lime glass apparatus calibrated at  $t_s$ , before entering Table 53 or Table A as appropriate. This always applies irrespective of the glass from which the pycnometer is made, i. e., enter Table 53 or Table A with  $\rho_t^1$  and against  $t_t$  extract  $\rho_{15}$  or  $\rho_{20}$  respectively, interpolating as necessary, where

a) for liquids

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [ 1 + \alpha_2 ( t_s - t_t ) ]$$

b) for solids or semi-solids

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o - m_2 + m_1) 1000} + C \right] [ 1 + \alpha_2 ( t_s - t_t ) ]$$

#### 17.4.2.3 When $t_c = t_s \neq t_t$ ( $t_s = 15^\circ$ or $20^\circ\text{C}$ )

a) Pycnometer made of soda-lime glass — Tables 53A, 53B, 53D or A may be entered directly with  $\rho_t^1$  and against  $t_t$  the corresponding  $\rho_{15}$  or  $\rho_{20}$  is extracted, interpolating as necessary, where

$$\rho_t^1 = \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C$$

b) Pycnometers made of borosilicate glass — Tables 53A, 53B, 53D or Table A must be entered with  $\rho_t^1$  adjusted for the differential expansion of the pycnometer glass, and against  $t_t$  the corresponding  $\rho_{15}$  or  $\rho_{20}$  is extracted, interpolating as necessary, where

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [ 1 + (\alpha_2 - \alpha_1) ( t_s - t_t ) ]$$

Second-order terms, which are not again significant, are ignored.

**17.4.2.4** When  $t_t \neq t_c \neq t_s$ 

The density determined at the test temperature  $t_t$  as in 17.4.1.2 is the true density at  $t_t$ . The result obtained must, therefore, be adjusted to give the apparent density at  $t_t$  in soda-lime glass apparatus calibrated at  $t_s$  before entering Tables 53A, 53B, 53D, or Table A as appropriate. This always applies, irrespective of the glass from which the pycnometer is made, i. e. enter Tables 53A, 53B, 53D, or Table A with  $\rho_t^1$  and against  $t_t$  extract  $\rho_{15}$  or  $\rho_{20}$  respectively, interpolating as appropriate. The expression for  $\rho_t^1$  may be rearranged as follows, ignoring second-order terms which are not significant:

a) Pycnometers made of soda-lime glass

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [1 + \alpha_2 (t_s - t_c)]$$

b) Pynometers made of borosilicate glass

$$\rho_t^1 = \left[ \frac{(m_t - m_o) \rho_c}{(m_c - m_o) 1000} + C \right] [1 + \alpha_1 t_s + \alpha_2 t_t - (\alpha_1 + \alpha_2) t_c]$$

**17.4.2.5** When  $t_s = t_t \neq t_c$ 

Use the expression in 17.4.1.2.

**17.4.3** *Rounding-off of Calculations and Reporting*

Carry out all calculations to five decimal places and report the final result to the nearest 0.000 1 as Density, giving the units of mass and volume and the temperature.

**17.5** Calculation of Relative Density

**17.5.1** In accordance with the definition, the appropriate values for  $d_t$ ,  $d_t^1$  and  $d_s$  can be derived by dividing the corresponding values for density in 17.4.1 and 17.4.2 by the density of water in the same units at the required reference temperature (see Note under 17.5.3).

**17.5.2** When the correction of relative density  $t/(60^\circ\text{F}) 15.6^\circ\text{C}$  of petroleum and petroleum products is required to the reference temperature of  $(60^\circ\text{F}) 15.6^\circ\text{C}$  (see 17.4.2.2, 17.4.2.3 and 17.4.2.4), use Tables 23A, 23B or 23D or IP 200  $t_s$ ,  $t_t$  and  $t_c$  will be in  $^\circ\text{F}$  and the coefficient of cubical expansion  $\alpha_1$  and  $\alpha_2$  should be stated per  $1^\circ\text{F}$  for  $\alpha_2$ , the value of  $14 \times 10^{-6}$  per  $1^\circ\text{F}$  may be taken.

**17.5.3** Carry out all calculations to five decimal places and report the final results to the nearest 0.000 1 as Relative Density giving test and reference temperatures.

NOTE — Since the densities in Table 3 are given in  $\text{kg/m}^3$ , it is essential that these values be divided by 1 000 to convert them to  $\text{g/ml}$  ( $\text{kg/l}$ ).

**18** PRECISION

**18.1** With samples which are neither very volatile nor very viscous, and using a pycnometer of at least 25 ml capacity, results of duplicate test should not differ by more than the following amounts:

	Repeatability	Reproducibility
Density	0.000 6 g/ml	0.000 6 g/ml
Relative density	0.000 6 g/ml	0.000 6 g/ml

These values are based on estimates of normal good practice and are not derived by statistical analysis.

For volatile or very viscous samples and for solids no specific figures for precision can be given.

**19** METHOD D (DENSITY OR RELATIVE DENSITY OF SOLIDS — DISPLACEMENT METHOD)**19.1** Outline of the Method

When a solid is weighed first in air and then in water, the difference between the two weighings is equal to the weight of water displaced. The rel.  $d$   $t/t$  of the solid may, therefore, be calculated, where  $t$  is the temperature of the water. If the solid has a relative density greater than unity a piece of it may be weighed directly in the water. If the relative density is less than unity the solid must be made to sink in the water in some suitable way, for example, by securing it to a metal crucible.

The method may be used for the determination of the relative density of bitumen and similar products which are sufficiently solid to be handled in pieces. The pycnometer method is, however, more precise.

**20** APPARATUS

**20.1** Thermometer — See 3.4.1.1.

**20.2** Balance Straddle — a pan straddle of convenient size to support a beaker and permit determination of the weight of the specimen in water.

**20.3** Waxed Thread — a length of fine, waxed silk thread.

**20.4 Nickel Crucible** — in a modification of the method a thin nickel crucible of 5-25 ml capacity may be used. It should be provided with a three-wire suspension which permits the complete immersion of wire and crucible when determining the weight in water.

## 21 PROCEDURE FOR SOLIDS OF RELATIVE DENSITY GREATER THAN UNITY

**21.1** Suspend a piece of the material, as smooth as possible, free from cracks and dust, and weighing at least 5 g, from the hook on one side of a balance, using a suitable length of waxed silk thread so that the bottom of the specimen is about 25.4 mm (1 in) above the straddle. Tare the thread previously by a length of thread on the other pan. Weigh the sample to the nearest mg. Place the sample in a beaker of freshly boiled distilled water maintained at the desired temperature, e. g. 15°C, and allow to remain for sufficient time to ensure that the sample has attained that temperature. Then place the beaker containing distilled water on the pan straddle and suspend the sample from the hook on one side of the balance. Remove with a fine wire any adhering air bubbles and determine the weight of the sample in water.

NOTE — The main source of error is air bubbles either occluded in the sample or adhering to its outer surface during weighing in water. Errors caused by weighing or by temperature variation are usually small in comparison.

## 22 CALCULATION AND REPORTING

**22.1** If  $a$  is the weight of the specimen in air, and  $b$  is the weight of the specimen in water at  $t$ , then

$$\text{relative } dt/t \text{ in air} = \frac{a}{a - b}$$

**22.2** Report the value so obtained, to the nearest 0.001, as Relative Density  $dt/t$ , Method D. The buoyancy correction in this case is negligible.

**22.3** To obtain density multiply the relative density  $dt/t$  by the density of water at  $t$ , obtained from Table 3.

## 23 PROCEDURE FOR SOLIDS OF RELATIVE DENSITY LESS THAN UNITY

In using the crucible method:

**23.1** Suspend the crucible from the hook on one side of the balance by means of a tared length of waxed thread, and determine its weight to the nearest mg.

**23.2** Suspend the crucible in a beaker containing freshly boiled distilled water at the desired

temperature so that the crucible and wire suspension are completely immersed; remove any air bubbles adhering to the crucible and determine the weight in water.

**23.3** Dry the crucible and pour in some of the molten sample, taking care to avoid inclusion of air bubbles, after which allow the crucible and contents to cool, preferably under vacuum to room temperature. Determine the weight in air.

**23.4** Place the crucible and content in a beaker of freshly boiled distilled water maintained at the desired temperature and leave until the crucible and contents attain that temperature.

**23.5** Place the beaker containing distilled water on the pan straddle and suspend the crucible from the hook on one side of the balance. Remove with a fine wire any air bubbles adhering to the crucible or sample, and determine the weight in water of the crucible and contents.

## 24 CALCULATION AND REPORTING

**24.1** Let  $W_c$  = weight of crucible in air;

$W_{cw}$  = weight of crucible in water at  $t$ ;

$W_f$  = weight of crucible + sample in air;

$W_{fw}$  = weight of crucible + sample in water at  $t$ ;

then  $W_f - W_c$  = weight of sample in air;

and  $(W_f - W_{cw}) - (W_{fw} - W_{cw})$

= loss in weight of sample when weighed in water at  $t$ .

therefore,

relative density  $dt^\circ\text{C}/t^\circ\text{C}$  in air

$$= \frac{W_f - W_c}{(W_f - W_c) - (W_{fw} - W_{cw})}$$

**24.2** Report the result as Relative Density, Method D. The buoyancy correction in this case is negligible.

**24.3** To obtain density multiply the relative density  $dt/t$  by the density of water at  $t$ , obtained from Table 3.

## 25 PRECISION

Results of duplicate tests should not differ by more than the following amounts:

Size of Sample	Repeatability	Reproducibility
5 g or larger	0.004	0.008
Less than 5 g	No specific permissible variations can be given	

Table 3 Density of Water in kg/m<sup>3</sup> Against °C of the International Practical Scale of Temperature ( 1968 )

( Clause 9.1 )

$t_{68}^{\circ}\text{C}$	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9	$\frac{\Delta\rho}{\Delta t}$ kg/m <sup>3</sup> 0.1°C	$\Delta\rho$ Air- Saturated Minus Air-Free
0	999.839 6	999.846 3	999.852 8	999.859 1	999.865 3	999.871 3	999.877 1	999.882 7	999.888 2	999.893 4	0.005 9	-0.002 6
1	999.898 5	999.903 5	999.908 2	999.912 8	999.917 2	999.921 4	999.925 4	999.929 3	999.933 0	999.936 5	0.004 1	-0.002 7
2	999.939 9	999.943 1	999.946 1	999.948 9	999.951 6	999.954 1	999.956 5	999.958 7	999.960 7	999.962 5	0.002 4	-0.002 8
3	999.964 2	999.965 7	999.967 0	999.968 2	999.969 2	999.970 1	999.970 8	999.971 3	999.971 7	999.971 9	0.000 8	-0.003 0
4	999.972 0	999.971 8	999.971 6	999.971 1	999.970 5	999.969 8	999.968 9	999.967 8	999.966 6	999.965 2	-0.000 8	-0.003 1
5	999.963 7	999.962 0	999.960 2	999.958 2	999.956 0	999.953 7	999.951 3	999.948 7	999.945 9	999.943 0	-0.002 4	-0.003 3
6	999.939 9	999.936 7	999.933 4	999.929 9	999.926 2	999.922 4	999.918 4	999.914 3	999.910 1	999.905 7	-0.003 9	-0.003 4
7	999.901 1	999.896 4	999.891 6	999.886 6	999.881 5	999.876 2	999.870 8	999.865 2	999.859 5	999.853 7	-0.005 3	-0.003 5
8	999.847 7	999.841 6	999.835 3	999.828 9	999.822 3	999.815 7	999.808 8	999.801 9	999.794 7	999.787 5	-0.006 8	-0.003 5
9	999.780 1	999.772 6	999.764 9	999.757 1	999.749 2	999.741 1	999.732 9	999.724 6	999.716 1	999.707 5	-0.008 1	-0.003 4
10	999.698 7	999.689 8	999.680 8	999.671 7	999.662 4	999.653 0	999.643 4	999.633 7	999.623 9	999.614 0	-0.009 5	-0.003 3
11	999.603 9	999.593 7	999.583 4	999.572 9	999.562 3	999.551 6	999.540 8	999.529 8	999.518 7	999.507 4	-0.010 8	-0.003 1
12	999.496 1	999.484 6	999.473 0	999.461 2	999.449 4	999.437 4	999.425 3	999.413 0	999.400 7	999.388 2	-0.012 1	-0.002 9
13	999.375 6	999.362 8	999.350 0	999.337 0	999.323 9	999.310 6	999.297 3	999.283 8	999.270 2	999.256 5	-0.013 3	-0.002 6
14	999.242 7	999.228 7	999.214 6	999.200 4	999.186 1	999.171 7	999.157 1	999.142 4	999.127 6	999.112 7	-0.014 5	-0.002 3
15	999.097 7	999.082 6	999.067 3	999.051 9	999.036 4	999.020 8	999.005 1	998.989 2	998.973 3	998.957 2	-0.015 7	-0.002 0
16	998.941 0	998.924 7	998.908 3	998.891 7	998.875 1	998.858 3	998.841 4	998.824 4	998.807 3	998.790 1	-0.016 8	-0.001 7
17	998.772 8	998.755 3	998.737 8	998.720 1	998.702 3	998.684 5	998.666 5	998.648 3	998.630 1	998.611 8	-0.017 9	-0.001 4
18	998.593 4	998.574 8	998.556 2	998.537 4	998.518 5	998.499 5	998.480 4	998.461 2	998.441 9	998.422 5	-0.019 0	-0.001 1
19	998.403 0	998.383 3	998.363 6	998.343 8	998.323 3	998.303 7	998.283 6	998.263 3	998.242 9	998.222 4	-0.020 1	-0.000 9
20	998.201 9	998.181 2	998.160 4	998.139 5	998.118 5	998.097 3	998.076 1	998.054 8	998.033 4	998.011 9	-0.021 2	-0.000 6
21	997.990 2	997.968 5	997.946 7	997.924 7	997.902 7	997.880 5	997.858 3	997.836 0	997.813 5	997.791 0	-0.022 2	-0.000 4
22	997.768 3	997.745 6	997.722 7	997.699 8	997.676 7	997.653 6	997.630 3	997.607 0	997.583 5	997.560 0	-0.023 2	-0.000 2
23	997.536 3	997.512 6	997.488 7	997.464 8	997.440 8	997.416 6	997.392 4	997.368 0	997.343 6	997.319 1	-0.024 2	-0.000 1
24	997.294 4	997.269 7	997.244 9	997.220 0	997.195 0	997.169 9	997.144 6	997.119 3	997.093 9	997.068 5	-0.025 2	-0.000 0
25	997.042 9	997.017 2	996.991 4	996.965 5	996.939 6	996.913 5	996.887 3	996.861 1	996.834 7	996.808 3	-0.026 1	-0.000 0
26	996.781 8	996.755 1	996.728 4	996.701 6	996.674 7	996.647 7	996.620 6	996.593 4	996.566 1	996.538 8	-0.027 0	-0.000 0
27	996.511 3	996.483 7	996.456 1	996.428 4	996.400 5	996.372 6	996.344 6	996.316 5	996.288 3	996.260 0	-0.028 0	-0.000 0
28	996.231 6	996.203 2	996.174 6	996.146 0	996.117 2	996.088 4	996.059 5	996.030 5	996.001 4	995.972 2	-0.028 9	-0.000 0
29	995.943 0	995.913 6	995.884 2	995.854 6	995.825 0	995.795 3	995.765 5	995.735 6	995.705 6	995.675 6	-0.029 8	-0.000 0
30	995.645 4	995.615 2	995.584 8	995.554 4	995.523 9	995.493 4	995.462 7	995.431 9	995.401 1	995.370 1	-0.030 6	0.000 0
31	995.339 1	995.308 0	995.276 8	995.245 6	995.214 2	995.182 8	995.151 2	995.119 6	995.087 9	995.056 1	-0.031 5	0.000 0
32	995.024 3	994.992 3	994.960 3	994.928 2	994.896 0	994.863 7	994.831 3	994.798 3	994.766 3	994.733 7	-0.032 3	0.000 0
33	994.701 0	994.668 2	994.635 3	994.602 1	994.569 3	994.536 2	994.503 0	994.469 7	994.436 4	994.402 9	-0.033 2	0.000 0
34	994.369 4	994.335 8	994.302 1	994.268 3	994.234 5	994.200 5	994.166 5	994.132 4	994.098 2	994.064 0	-0.034 0	0.000 0
35	994.029 6	993.995 2	993.960 7	993.926 1	993.891 5	993.856 7	993.821 9	993.787 0	993.752 1	993.717 0	-0.034 8	0.000 0
36	993.681 9	993.646 7	993.611 4	993.576 0	993.540 6	993.505 0	993.469 4	993.433 8	993.398 0	993.362 2	-0.035 6	0.000 0
37	993.326 3	993.290 3	993.254 2	993.218 1	993.181 8	993.145 5	993.109 2	993.072 7	993.036 2	993.999 6	-0.036 3	0.000 0
38	992.962 9	992.926 1	992.889 8	992.852 4	992.815 4	992.778 4	992.741 2	992.704 0	992.666 8	992.629 4	-0.037 1	0.000 0
39	992.592 0	992.554 5	992.516 9	992.479 2	992.441 5	992.403 7	992.365 8	992.327 9	992.289 9	992.251 8	-0.037 8	0.000 0
40	992.213 6											

Table 4 Buoyancy Corrections\*

( Clause 17.1 )

$\frac{m_t - m_o}{M}$	Correction, g/ml	$\frac{m_t - m_o}{M}$	Correction, g/ml	$\frac{m_t - m_o}{M}$	Correction, g/ml
0.60	0.000 48	0.80	0.000 24	1.00	0.000 00
0.61	0.000 47	0.81	0.000 23	1.01	— 0.000 01
0.62	0.000 46	0.82	0.000 22	1.02	— 0.000 02
0.63	0.000 44	0.83	0.000 20	1.03	— 0.000 04
0.64	0.000 43	0.84	0.000 19	1.04	— 0.000 05
0.65	0.000 42	0.85	0.000 18	1.05	— 0.000 06
0.66	0.000 41	0.86	0.000 17	1.06	— 0.000 07
0.67	0.000 40	0.87	0.000 16	1.07	— 0.000 08
0.68	0.000 38	0.88	0.000 14	1.08	— 0.000 10
0.69	0.000 37	0.89	0.000 13	1.09	— 0.000 11
0.70	0.000 36	0.90	0.000 12	1.10	— 0.000 12
0.71	0.000 35	0.91	0.000 11	1.11	— 0.000 13
0.72	0.000 34	0.92	0.000 10	1.12	— 0.000 14
0.73	0.000 32	0.93	0.000 08	1.13	— 0.000 16
0.74	0.000 31	0.94	0.000 07	1.14	— 0.000 17
0.75	0.000 30	0.95	0.000 06	1.15	— 0.000 18
0.76	0.000 29	0.96	0.000 05	1.16	— 0.000 19
0.77	0.000 28	0.97	0.000 04	1.17	— 0.000 20
0.78	0.000 26	0.98	0.000 02	1.18	— 0.000 22
0.79	0.000 25	0.99	0.000 01	1.19	— 0.000 23

\*Calculated for standard air of density 0.001 222 g/ml at 15.56°C ( 60°F ) and at a pressure of 1 : 013 bar ( 760 mm mercury ).

These corrections are applicable for air density values between 0.001 1 and 0.001 3 g/ml.

NOTE —  $M$  = Weight in air water contained by the pyknometer at temperature  $t$ .



### **Standard Mark**

The use of the Standard Mark is governed by the provisions of the *Bureau of Indian Standards Act, 1986* and the Rules and Regulations made thereunder. The Standard Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well defined system of inspection, testing and quality control which is devised and supervised by BIS and operated by the producer. Standard marked products are also continuously checked by BIS for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.

## Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 1986* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

### Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

### Revision of Indian Standards

Indian Standards are reviewed periodically and revised, when necessary and amendments, if any, are issued from time to time. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition. Comments on this Indian Standard may be sent to BIS giving the following reference:

Doc : No. PCD 1 ( 1086 )

### Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

## BUREAU OF INDIAN STANDARDS

### Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110002  
Telephones : 331 01 31, 331 13 75

Telegrams : Manaksanstha  
( Common to all Offices )

### Regional Offices :

Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg  
NEW DELHI 110002

Telephone

{ 331 01 31  
{ 331 13 75

Eastern : 1/14 C. I. T. Scheme VII M, V. I. P. Road, Maniktola  
CALCUTTA 700054

{ 37 84 99, 37 85 61,  
{ 37 86 26, 37 86 62

Northern : SCO 445-446, Sector 35-C, CHANDIGARH 160036

{ 53 38 43, 53 16 40,  
{ 53 23 84

Southern : C. I. T. Campus, IV Cross Road, MADRAS 600113

{ 235 02 16, 235 04 42,  
{ 235 15 19, 235 23 15

Western : Manakalaya, E9 MIDC, Marol, Andheri ( East )  
BOMBAY 400093

{ 632 92 95, 632 78 58,  
{ 632 78 91, 632 78 92

Branches : AHMADABAD. BANGALORE. BHOPAL. BHUBANESHWAR. COIMBATORE.  
FARIDABAD. GHAZIABAD. GUWAHATI. HYDERABAD. JAIPUR. KANPUR.  
LUCKNOW. PATNA. THIRUVANANTHAPURAM.